

User manual
KP-20 Series

(KP-20.2 · KP-25.2)



UK
CA

GENERAL WARNING

IMPORTANT

Do not attempt to use the wheelchair before a qualified person has explained to your satisfaction how to operate it without risk to yourself or others.



WARNING

- Do not operate this powered wheelchair before first reading and understanding this owner's manual. If you are unable to understand the warnings, cautions, suggestions and technical instructions, please contact a dealer, healthcare professional or applicable personnel before attempting to use this power wheelchair -- otherwise, injury or damage may occur. If any procedures other than those described in this owner's manual, it **MUST** be performed by a qualified technician.

THE INFORMATION CONTAINED IN THIS DOCUMENT IS SUBJECT TO CHANGE WITHOUT NOTICE BY KARMA.



WARNING

- As a manufacturer of powered wheelchairs, KARMA endeavors to supply a wide variety of powered wheelchairs to meet the many needs of the user. However, the responsibility of final selection of the type of powered wheelchair to be used by an individual rests solely with the user and his/her healthcare professional capable of making such a selection. KARMA strongly recommends consulting a certified rehab technology supplier or assistive technology professional.

It is the responsibility of the Durable Medical Equipment (DME) Dealer, therapists and other healthcare professionals to determine if a seat positioning strap is required to ensure the safe operation of this equipment by the user. Serious injury can occur in the event of a fall from a powered wheelchair.

Operating Information

- Performance adjustments should ONLY be made by professionals of the healthcare field or persons fully conversant with this process and the driver's capabilities. Incorrect settings could cause injury to the driver or bystanders and damage to the powered wheelchair and surrounding property.
- To determine and establish your particular safety limits, practice bending, reaching, mounting and dismounting activities in several combinations in the presence of a qualified healthcare professional BEFORE attempting active use of the powered wheelchair.
- For individuals with balance problems, practice mounting and dismounting activities WITH AN ASSISTANT in the presence of a qualified healthcare professional.
- If so equipped, DO NOT operate the wheelchair without anti-tippers being installed. Anti-tippers MUST BE attached at all times
- DO NOT attempt to reach objects if you have to move forward in the seat or pick them up from the floor by reaching down between your knees as this may cause loss of balance.
- DO NOT lean over the back of the chair to reach objects behind you as this may cause the powered wheelchair to tip over.
- DO NOT shift your weight or sitting position toward the direction you are reaching as the powered wheelchair may tip over.
- DO NOT use an escalator to move a powered wheelchair between floors. Serious bodily injury may occur.
- DO NOT operate in car lanes on roads, streets, highways, etc.
- DO NOT attempt to move up or down an incline with a water, ice or oil film.
- DO NOT attempt to ride over curbs or obstacles. Doing so may cause your powered wheelchair to turn over and cause bodily harm and/or damage to the powered wheelchair.
- DO NOT make sharp turns, in forward or reverse, at excessive speeds.
- DO NOT attempt to lift the powered wheelchair by its casings and seat.
- DO NOT attempt to lift the powered wheelchair by any detachable parts. Lifting by means of any detachable parts of a powered wheelchair may result in injury to the user or damage to the powered wheelchair. Please hold onto the frame when lifting your wheelchair.
- DO NOT operate the powered wheelchair until you have checked that the surroundings are clear and that the area is safe for travel.
- Please pay attention to the environment, and to keep away from the hazard of external flame. When the user is smoking, pay attention to safety and make sure to extinguish the fire when finished. Be alert to the environment when someone nearby is smoking or using a cigarette lighter, keep away from the hazard and use fire-proof cloth.
- DO NOT use parts, accessories or adapters other than those authorized by KARMA. Before attempting to sit in or exit the power wheelchair, turn the power OFF. This will ensure that the power wheelchair will not drive. Turn the power OFF while the

wheelchair is not in use. Otherwise, injury and/or damage to the wheelchair and surrounding property may occur

- DO NOT connect any medical device (i.e., a ventilator, life support machine, et cetera) to the battery. This could cause unexpected failure of the device and the powered wheelchair. KARMA specifically disclaims responsibility to all personal injury and property damage that may occur during use which does not comply with applicable federal, state and local laws and ordinances.
- DO NOT stand on the footplates of the powered wheelchair.
- DO NOT use the wheelchair when the back is folded. This alters the centre of gravity and may cause the wheelchair to tip backwards. The back should only be folded for storage or transport of the wheelchair.
- DO NOT hang anything on the backjoints that could cause the backrest folding suddenly and may cause injury.
- DO NOT be towed by other vehicles(cars, motorcycles or bicycles, etc.). Only use freewheel mode on flat surfaces since your electromagnetic brakes are not engaged in this mode, without these brakes it will roll down a hill causing injury or damage. Never drive under the influence of alcohol or medication.
- **DO NOT** sit in the powered wheelchair within a moving vehicle unless it has been affixed the labels of hook mark. The labels of hook mark means that the powered wheelchair fulfils the performance requirements for a "Crash Test" of ISO 7176-19 and is suitable to use as a seat in vehicles that have been adapted for this purpose.



WARNING

- **Please transfer to a seat in the car if at all possible.**

Ramps

- DO NOT climb, go up or down ramps or traverse slopes greater than the Max Safe Slope for your model specified in Section 10.
- When negotiating ramps, if the joystick is released while moving forward, the powered wheelchair will roll backwards approximately one foot before the brake engages. If the joystick is released while in reverse, the powered wheelchair will roll backwards approximately two and a half feet before the brake engages.
- Be sure to reduce your speed when driving on curved roads or making a turn; do not drive at full speed. Please drive at a slower speed so you have complete control of the wheelchair.

Electrical

- Check to ensure that all electrical connections are secure at all times.
- Grounding Instructions: DO NOT, under any circumstances, cut or remove the round grounding prong from any plug. Some devices are equipped with three-prong (grounding) plugs for protection against possible shock hazards.
- Where a two-prong wall receptacle is encountered, it is the personal responsibility and obligation of the customer to contact a qualified electrician and have the

two-prong receptacle replaced with a properly grounded three-prong wall receptacle/outlet in accordance with the National Electrical Code and local laws.

- Do not use a n extension cord when charging your batteries. A risk of fire and /or electric shock could be encountered.

Batteries

- The warranty and performance specifications contained in this manual are based on the use of deep cycle sealed lead acid batteries.
- KARMA strongly recommends their use as the power source for this unit. (Refer to Section 6)
- Carefully read the battery and charger information prior to installing, servicing or operating your wheelchair.

Rain Test

- DO NOT leave your powered wheelchair in a rain storm of any kind.
- DO NOT use your powered wheelchair in a shower or leave it in a damp bathroom while taking a shower.
- DO NOT leave your powered wheelchair in a damp area for any length of time. Direct exposure to rain or dampness could cause the powered wheelchair to malfunction electrically and mechanically and may cause the powered wheelchair to prematurely rust.

Stairways

- DO NOT attempt to move an occupied powered wheelchair between floors using a stairway. **(See Figure 1)**
- Use an elevator to move an occupied powered wheelchair between floors. If moving a powered wheelchair between floors by means of a stairway, the occupant **MUST** be removed and transported independently of the powered wheelchair. Extreme caution is advised when it is necessary to move an **UNOCCUPIED** powered wheelchair up or down stairs.
- KARMA recommends disassembling the wheelchair and transporting the components independently up or down stairs. Make sure to use **ONLY** secure, non-detachable parts on each component for hand-hold supports.
- DO NOT attempt to lift a powered wheelchair by any removable (detachable) parts. Lifting by means of removable (detachable) parts may result in injury to the user or assistants or damage to the wheelchair.



(Figure 1)

Escalators

- DO NOT use an escalator to move a powered wheelchair between floors. Serious bodily injury and damage to the wheelchair may occur.

You and Your KARMA Service Provider

- No tools are required for assembly. Your wheelchair should be assembled by your authorized KARMA dealer. Your authorized KARMA dealer is also your service provider. Unauthorized repairs, upgrades and add-ons will negate your warranty.
- Anything wheelchair related that requires tools should be done by your authorized KARMA service provider. Your KARMA service provider has the necessary service manual for your wheelchair.
- If you have any defective parts or lose any parts, please contact your dealer for repair and/or replacement. Please only use parts authorized by KARMA. Use of parts that are not authorized by KARMA will negate your warranty.
- Part numbers are listed on our bill of materials (explosion drawings), which we give to our authorized dealers and service providers. Customers should order parts through their dealer as well as all repairs and scheduled maintenance.
- Customers should go to their dealership for all service needs to guarantee their warranty is not void. Always go to an authorized service provider. Unauthorized repairs will void your warranty. Procedures other than those described in this manual must be performed by a qualified technician. They also have the service manual with a complete list of maintenance required to keep your KARMA wheelchair in top performance.
- Check the warranty section of this manual to find out what KARMA guarantees and what maintenance service and parts that are not covered by the warranty.
- KARMA does not have an official list of service providers at this time. Please contact your dealer for service. If you cannot contact your dealer, please contact another authorized KARMA dealer for service. If all else fails, contact KARMA directly via our website or give us a call and we can refer you to your new service provider.

CLASSIFICATION & INTENDED USE

- KARMA's KP-20.2/KP-25.2 are classified as a class B mobility product for indoor and outdoor use. This wheelchair is suitable for users up to 135kg/115kg(w/solid wheel) in weight. The KP-20.2/KP-25.2 in Australia was tested for 125kgs conforms to AU/NZS. It also has an extremely durable aluminum alloy frame with tough accessory and does not suffer in dramatic extremes of temperature (neither user nor attendant need touch metal parts in normal use).
- Check with your local government as to legislation pertaining to mobility vehicles to assure your legal requirements to drive on public roads have been met, i.e. a pelvic belt, lighting, registration or licensing.

CE MARKING

- This powered wheelchair complies with the requirements of the Medical Devices Regulation (EU) 2017/745.



CAUTION

- Before purchasing and using KARMA wheelchairs, please consult qualified professionals so as to ensure that you choose suitable products and use them correctly as well. For safety reasons, if there is any risk due to the user's movement in the wheelchair or other environmental factors that might cause the wheelchair to tip or the person on the wheelchair to fall off, KARMA recommends that you install anti-tippers, pelvic belt, or other additional safety accessories.



WARNING

- When prescribing wheelchairs for use by full or partial amputees (above or below knee, single or double) or other conditions that affects the user's natural centre of gravity and might cause the wheelchair to tip or the person on the wheelchair to fall off, we suggest contacting your therapist or dealer as well as attaching anti-tippers, pelvic belt, safety belt, or other additional safety accessories.
- When using the wheelchair, ALWAYS follow the Highway Code or your local traffic regulations when outdoors and the guidelines written in this manual.

Declaration of conformity



Document No.: DOC-KP-2-02

EU Declaration of Conformity

We, with the information specified in the below,

Our information as the manufacturer:

Manufacturer's Name:	KARMA MEDICAL PRODUCTS CO., LTD
Manufacturer's Address:	NO. 2363, Sec. 2, University Rd., Min-Hsiung Shiang, Chia-Yi County, 62144, Taiwan
SRN (Single Registration Number):	TW-MF-000013206

Our authorized representative:

Name:	KARMA MOBILITY, S.L.
SRN (Single Registration Number):	ES-AR-000004852
Address:	C/ PERIODISTA FRANCISCO CARANTOÑA DUBERT, 23 Bajo 33209 GIJÓN – ASTURIAS, SPAIN
Contact Person:	Raquel Yuste
Contact Information:	(+34) 984 390 907

in accordance with

Regulation (EU) 2017/745 of the European Parliament and of the council Annex I, II, III, IV and IX

hereby declare that the medical device specified below:

Basic UDI-DI of Annex VI:	471987385KP-2LA
Device:	Electrically powered wheelchair
Trade Name or Mark:	Falcon / KP-25.2 series
Model Number:	KP-20.2, KP25.2
Product Code according to EMDN:	Y122127
Product Code according to GMDN:	41637
Classification:	Rule 1 of Class I
UDI-DI:	04719873856640 04719873856657

is in conformity with the applicable requirements of the following documents:

Ref. No.	Title	Edition date
ISO 13485	Medical devices - Quality management systems - Requirements for regulatory purposes	2016
ISO 14971	Medical devices - Application of risk management to medical devices	2019
EN 12182	Assistive products for persons with disability - General requirements and test methods	2012
EN 12184	Electrically powered wheelchairs, scooters and their chargers - Requirements and test methods	2014
EN 62366	Medical devices - Application of usability engineering to medical devices	2015
EN 60601-1	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance	2006
EN ISO 10993-1	Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process	2009
EN ISO 10993-5	Biological evaluation of medical devices — Part 5: Tests for in vitro cytotoxicity	2009
EN ISO 15223-1	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General requirements	2016
MEDDEV. 2.7/1 Rev. 4	CLINICAL EVALUATION: A GUIDE FOR MANUFACTURERS	2016
MEDDEV 2.12/1 Rev. 8	GUIDELINES ON A MEDICAL DEVICES VIGILANCE SYSTEM	2013

The information on this declaration has been stated on the sole responsibility of KARMA MEDICAL PRODUCTS CO., LTD.

We hereby declare that the device named above has been designed to comply with the relevant sections of the above referenced specifications. The device complies with all General Safety and Performance Requirements.

Date of issue: 23th August 2022

Place of issue: NO. 2363, Sec. 2, University Rd., Min-Hsiung Shiang, Chia-Yi County, 62144,
Taiwan


Richard Chang, CEO

UK Declaration of Conformity

We, with the information specified in the below,

Our information as the manufacturer:

Manufacturer's Name:	KARMA MEDICAL PRODUCTS CO., LTD
Manufacturer's Address:	NO. 2363, Sec. 2, University Rd., Min-Hsiung Shiang, Chia-Yi County, 62144, Taiwan
SRN (Single Registration Number):	TW-MF-000013206

Our UK responsible person:

Name:	Karma Mobility Ltd
MHRA Reference Number:	11480
Address:	Unit 6 Target Park, Shawbank Road, Redditch, B98 8YN England United Kingdom
Contact Person:	Mr Mark Duffield
Contact Information:	01527 520 468

in accordance with

UK MDR 2002 and Council Directive 93/42/EEC concerning medical devices in Annex I, II and VII hereby declare that the medical device specified below:

Basic UDI-DI of MDR(Annex VI):	471987385KP-2LA
Device:	Electrically powered wheelchair
Trade Name or Mark:	Falcon / KP-25.2 series
Model Number:	KP-25.2
Product Code according to EMDN:	Y122127
Product Code according to GMDN:	41637
Classification:	Rule 1 of Class 1
UDI-DI:	04719873856640 04719873856657

is in conformity with the applicable requirements of the following documents:

Ref. No.	Title	Edition date
EN ISO 13485	Medical devices - Quality management systems - Requirements for regulatory purposes	2016
EN ISO 14971	Medical devices - Application of risk management to medical devices	2019
EN 12182	Assistive products for persons with disability - General requirements and test methods	2012
EN 12184	Electrically powered wheelchairs, scooters and their chargers - Requirements and test methods	2014
EN 62366	Medical devices - Application of usability engineering to medical devices	2015
EN 60601-1	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance	2006
EN ISO 10993-1	Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process	2009
EN ISO 10993-5	Biological evaluation of medical devices — Part 5: Tests for in vitro cytotoxicity	2009
EN ISO 15223-1	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General requirements	2016

The information on this declaration has been stated on the sole responsibility of KARMA MEDICAL PRODUCTS CO., LTD.

We hereby declare that the device named above has been designed to comply with the relevant sections of the above referenced specifications. The device complies with all Essential Requirements.

Date of issue: 23th August 2022

Place of issue: NO. 2363, Sec. 2, University Rd., Min-Hsiung Shiang, Chia-Yi County, 62144, Taiwan



Richard Chang, CEO

How to contact KARMA



Head office

KARMA Medical

Address: NO.2363, Sec. 2, University Road, Minxiong Township,
Chiayi County 621015, Taiwan (R.O.C.)

Phone: +886 5 206 6688

Email: globalsales@karma.com.tw

Website: www.karmamedical.com

European representative

KARMA Mobility S.L.

Address: C/ Periodista Francisco Carantoña Dubert,
23 Bajo 33209 Gijón - Asturias (Spain)

Phone: +34 984 390 907

Email: karma@karmamobility.es

Website: www.karmamobility.es

UK representative

KARMA Mobility Ltd.

Address: Unit 6 Target Park, Shawbank Road, Reddtich,
B98 8YN England United Kingdom

Phone: +44 845 630 3436

Email: info@karmamobility.co.uk

Website: www.karmamobility.co.uk



Produced and published by KARMA Medical, Taiwan.

Technical changes and print errors are prohibited.

TABLE OF CONTENTS

1. PREFACE.....	1
2. SAFETY.....	2
2.1 BEFORE DRIVING.....	2
2.2 TRAFFIC RULES.....	2
2.3 PRACTICE DRIVING	2
2.4 NO PASSENGERS.....	3
2.5 NO HAULING HEAVY GOODS	3
2.6 RAIN.....	3
2.7 WHILE DRIVING	3
2.8 GENERAL WARNING	3
2.9 SURFACE TEMPERATURES.....	3
2.10 IGNITABILITY STANDARDS.....	3
2.11 RAILROAD CROSSING.....	4
2.12 CIRCUMSTANCES TO AVOID.....	4
2.13 MOBILE PHONES AND OTHER ELECTRIC EQUIPMENT.....	4
2.14 RAMPS, INCLINES AND DROPS.....	4
2.15 MAXIMUM USER WEIGHT LIMIT.....	6
2.16 SERVICE LIFE	6
2.17 KP-20.2/KP-25.2 LABELING.....	7
3. EMI/RFI.....	8
3.1 ELECTROMAGNETIC INTERFERENCE FROM RADIO WAVE SOURCES.....	8
3.2 THE SOURCES OF RADIATED EMI CAN BE BROADLY CLASSIFIED INTO THREE TYPES.....	9
3.3 POWERED VEHICLE ELECTROMAGNETIC INTERFERENCE (EMI)	10
3.4 WARNINGS.....	10
4. PARTS.....	12
5. OPERATION.....	14
5.1 VR II JOYSTICK CONTROLLER.....	14
5.2 HOW TO OPERATE YOUR WHEELCHAIR.....	14
5.3 SHARK CONTROLLER (OPTIONAL).....	16
5.4. POWER ON/OFF.....	17
5.5 SPEED ADJUSTMENT BUTTON.....	17
5.6 FORWARD, REVERSE AND BRAKING.....	17
5.7 DRIVING SPEED	18
5.8 TO STOP.....	18
5.9 HORN BUTTON	18
5.10 BATTERY INDICATOR.....	19
5.11 UNFOLDING THE CHAIR	20
5.12 FOLDING THE WHEELCHAIR.....	20
5.13 FOOTREST REMOVAL AND REFITTING	20
5.14 ARMREST FLIP-BACK & HEIGHT ADJUSTMENT.....	21
5.15 FREEWHEEL LEVERS	21
5.16 TO SELECT NEUTRAL FROM THE ENGAGED POSITION	21
5.17 TO ENGAGE THE MOTORS FROM THE NEUTRAL POSITION	21
5.18 ARMREST FLIP UP AND HEIGHT ADJUSTMENT	22

5.19 GETTING IN AND OUT	22
5.20 GETTING IN	22
5.21 GETTING OUT	23
5.22 MOTOR BRAKES.....	24
5.23 PARKING BRAKES - HAND OPERATED	24
5.24 INSTALL THE ANTI-TIPPERS.....	25
5.25 TRANSPORTATION	25
5.26 BEFORE USE.....	26
5.27 METHOD OF RESTRAINT	26
5.28 FITTING THE SEAT BELT.....	27
5.29 ADJUSTING THE LENGTH OF THE PELVIC BELT.....	28
5.30 HOW TO OPERATE IT	29
5.31 MAINTENANCE	29
6. BATTERIES AND CHARGER.....	30
6.1 CHARGING THE BATTERIES	30
6.2 CHARGE YOUR BATTERIES IF ANY OF THE FOLLOWING CONDITIONS OCCUR	30
6.3 BE SURE TO FOLLOW THE PROCEDURES LISTED BELOW PRECISELY	30
6.4 CHARGER.....	32
6.5 BATTERIES	33
6.6 CLEANING THE BATTERIES	34
7. INSPECTION AND MAINTENANCE	35
7.1 DAILY CHECK.....	35
7.2 REGULAR MAINTENANCE RECORD.....	36
7.3 TIRES	37
7.4 GENERAL MAINTENANCE.....	38
7.5 SUGGESTED MAINTENANCE PROCEDURES	39
7.6 TRANSPORTING AND STORING	39
8. OPTIONAL ACCESSORIES.....	40
8.1 CUSHION	40
8.2 CURB CLIMBER.....	40
9. TROUBLESHOOTING.....	42
9.1 VRII CONTROLLER, BATTERY & DIAGNOSTICS INDICATOR.....	42
10. SPECIFICATIONS	46
11. CLEANING & RECYCLING	48
11.1 CLEANING	48
11.2 POST-CONSUMER RECYCLING	48
12. WARRANTY	50
12.1 SERIAL NUMBER	50
12.2 CONTENTS AND DURATION OF WARRANTY.....	50
12.3 THE CLAIMS FROM THIS WARRANTY SHALL NOT ARISE FOR THE FOLLOWING REASONS.....	51
12.4 THE RESPONSIBILITIES OF THE USER	51
12.5 THE RESPONSIBILITIES OF THE DEALER	52
12.6 TRANSFER OF WARRANTY RIGHTS.....	52

1. PREFACE

Please carefully read this owner's manual before using the wheelchair. Improper use of the wheelchair could result in harm, injury or traffic accidents. Therefore, for safe and enjoyable use of the wheelchair, please read this owner's manual.

1.1 This owner's manual includes operation instructions for the aspects of the wheelchair, assembly instructions, and instructions on how to deal with possible accidents. This owner's manual is written for KARMA powered wheelchair(s): KP-20.2/KP-25.2.

1.2 The symbols used in this manual are explained below. Pay special attention to the parts marked with these symbols.

 WARNING	Improper use could lead the user to severe injury or death.
 CAUTION	Improper use could lead the user to severe injury and/or damage to your wheelchair.
 SUGGESTION	Follow these instructions to keep the wheelchair in good condition.

1.3 This manual includes the repair and maintenance chart as well as the Warranty. Please keep it in a safe place or with the wheelchair.

1.4 If someone else uses the wheelchair, make sure that you give him or her this owner's manual for his or her reference.

1.5 As designs change, some illustrations and pictures in this manual may not correspond to the vehicle that you purchased. We reserve the right to make design modifications without further notice.

2. SAFETY

2.1 Before Driving

- 2.1.1 The user must be familiar with the use and operation of this wheelchair before driving.
- 2.1.2 Therefore, please always keep these safety guidelines in mind.

2.2 Traffic Rules

- 2.2.1 Pedestrian traffic rules apply to this wheelchair. Please be safe.
- 2.2.2 Drive only on pedestrian areas such as the sidewalk. Never drive the wheelchair on the highway.
- 2.2.3 Be aware of other vehicles when crossing all roads.
- 2.2.4 Be extremely cautious when driving your wheelchair in heavy traffic or the shopping mall.
- 2.2.5 DO NOT drive your wheelchair when you feel tired or consume alcohol.
- 2.2.6 Please do not drive your wheelchair at night.
- 2.2.7 Please follow pedestrian traffic lights.

2.3 Practice Driving

- 2.3.1 Before you are familiar with the operation of your wheelchair please practice in a wide and open area, like a park.
- 2.3.2 To avoid falling off your wheelchair please practice all kinds of driving motions, such as accelerating, stopping, turning, reversing and going up and down ramps.
- 2.3.3 Please set the speed to medium when you first start practicing.
- 2.3.4 Be safe. Make sure someone accompanies you incase you need assistance.
- 2.3.5 Be sure you are able to control and operate your wheelchair easily and confidently before you move the speed higher.

2.4 No Passengers

KARMA wheelchair is limited to one driver. Do not carry passengers (including children) on your wheelchair.

2.5 No Hauling Heavy Goods

Do not use this wheelchair to carry or haul heavy goods. The maximum weight that can be carried, including goods, is stated beside "Max. User Weight" in section 10. SPECIFICATIONS.

2.6 Rain

This product is not guaranteed to be water resistant. Please do not drive in the rain, through puddles or spray water to clean this product. (See Figure 2.1)

2.7 While Driving

Please carry out daily inspections. Refer to section 7.1 DAILY CHECK.

2.8 General Warning

Do not lean over the side of the wheelchair as such an action may cause you to lose balance and fall.

Use caution to assure your clothing doesn't get tangled with the wheels.

2.9 Surface temperatures

Surface temperatures can increase if the product is exposed to direct sunlight.

2.10 Ignitability standards

In compliance with ignitability standards. A separate report for resistance to ignition of upholstered components may be available from the manufacturer or supplier if requested.

2.11 Railroad Crossing

Before crossing the tracks, please stop completely and look both ways.

Cross the tracks perpendicular to avoid your wheels getting stuck.

Do not drive at full speed over railroad tracks.

2.12 Circumstances to Avoid

2.12.1 Avoid roads with heavy traffic, mud, excessive gravel or bumps, snow and ice. These conditions may damage your wheelchair.

2.12.2 Avoid roads that are too narrow or by a canal/waterway without any fence/hedge.

2.12.3 Also avoid places where your wheels might get stuck, slip or not have traction.

2.12.4 Do not drive in a gale, at night or in rainy/snowy/foggy/misty weather. These conditions may cause your wheelchair to rust. (See Figure 2.2)

2.12.5 Do not drive in a S curve, do doughnuts or make sudden turns.

2.12.6 Do not take an escalator.

2.13 Mobile Phones and Other Electric Equipment

2.13.1 Do not use a mobile phone or other wireless communication devices while driving. (See Section 3. EMI/RFI)

2.13.2 Do not charge the mobile phone or other electric devices from your wheelchair's batteries.

2.14 Ramps, Inclines and Drops

2.14.1 Do not drive onto a steep ramp. Refer to "Max. Climbing Ability" in Section 10. SPECIFICATIONS for your wheelchair's maximum climbing angle.

2.14.2 When climbing up to an inclined road, please set the speed higher than the medium speed and drive carefully. (See Figure 2.3)

2.14.3 When going down an inclined road, set the speed to "1" and never use reverse.

2.14.4 Do not drive on a road with many bumps and holes close to each other,

such as potholes or washboard roads. (See Figure 2.4)

2.14.5 Do not make sudden turns when driving on gravel roads or ramps.

2.14.6 The maximum surmountable obstacle height is 5.5 cm for your wheelchair(with optional curb climber is up to 8 cm). Exceeding this obstacle height will damage your wheelchair and void your warranty. (See Figure 2.5)

2.14.7 When negotiating an obstacle slow down, put your seat into the full upright position and approach it head on (perpendicular).

2.14.8 Just before the obstacle, speed up until the entire wheelchair has conquered the obstacle. Negotiating an obstacle at an angle may cause your wheelchair to tip over. Don't do it!

2.14.9 When climbing down a curb, slow down just before you get to it and only speed up after the entire wheelchair has lowered to the road.

2.14.10 Again, descending obstacles should be done perpendicularly. (See Figure 2.6)

2.14.11 KARMA does not design wheelchairs to take jumps over or off obstacles. Doing so will void your warranty.



WARNING

- **Do not set the wheelchair in freewheel mode when on an incline or decline.**
- **If you breakdown on railroad tracks, first, check for an oncoming train. If a train is coming, get out of its way immediately! If no train is coming, set to freewheel mode and push the wheelchair off the tracks.**

2.15 Maximum User Weight Limit

2.15.1 Refer to "Max. User Weight" in Section 10. SPECIFICATIONS.

2.15.2 Loading exceeding maximum capacity can damage your wheelchair and cause malfunctions yielding a safety hazard. The warranty does not cover damage caused by improper operation of the wheelchair.



(Figure 2.1)



(Figure 2.2)



(Figure 2.3)



(Figure 2.4)



(Figure 2.5)




(Figure 2.6)

2.16 Service life



The expected service life of the product is **five years**, presuming that the product is used daily in accordance with the safety operation instructions, inspection and maintenance instructions, and intended use, which described in this manual.

2.17 KP-20.2/KP-25.2 Labeling

Please carefully read all the labeling on the wheelchair before driving it. Do not remove them. Protect them for future reference.



Back side



Karma Medical Products Co., Ltd. 2363, Sec. 2 University Rd., Min-Hsiung, Chia-Yi 621, Taiwan		
TYPE KP-20.2 V max. : 6 km/h max. : 10° 115 kg 135 kg	 SN MD	
Karma Medical Products Co., Ltd. 2363, Sec. 2 University Rd., Min-Hsiung, Chia-Yi 621, Taiwan		
TYPE KP-25.2 V max. : 6 km/h max. : 10° 115 kg 135 kg	 SN MD	
KARMA MOBILITY, S. L. C/ PERIODISTA FRANCISCO CARANTOÑA DUBERT, 23 Bajo 33209 GIJÓN - ASTURIAS (SPAIN)		
KARMA MOBILITY LTD. UNIT 6 TARGET PARK, SHAWBANK ROAD, REDDITCH, B98 8YN ENGLAND UNITED KINGDOM		

3. EMI/RFI

This section provides the user with basic information about the problems with EMI, known sources of EMI and protective measures either to reduce the possibility of exposure or to minimize the degree of exposure. This section also shows some conditions in which unexpected or erratic wheelchair movements may occur.



CAUTION

- It is very important that you read this information regarding the possible effects of electromagnetic interference on your electric KARMA wheelchair

3.1 Electromagnetic Interference From Radio Wave Sources

3.1.1 Powered vehicles may be susceptible to electromagnetic interference (EMI), which is interfering electromagnetic energy (EM) emitted from sources such as radio stations (Radio Frequency Interference), TV stations, amateur radio (HAM) transmitters, two-way radios, and cellular phones. The interference (from radio wave sources) can cause the powered vehicle to release its brakes, move by itself, or move in unintended directions. It can also permanently damage the powered vehicle's control system. The intensity of the interfering EM energy can be measured in volts per meter (V/m). Each powered vehicle can resist EMI up to a certain intensity. This amount of resistance is called its immunity level. The higher the immunity level, the greater the protection. At this time, current technology is capable of achieving at least a 20 V/m immunity level, which would provide useful protection from the more common sources of radiated EMI. This powered vehicle model, with no further modification, has an immunity level of 20 V/m without any

accessories.

3.1.2 There are a number of sources of relatively intense electromagnetic fields in our everyday environment. Some of these sources are obvious and easy to avoid. Others are not apparent and exposure could be unavoidable. However, we believe that by following the warnings listed below, your risk to EMI can be greatly minimized.

3.2 The sources of radiated EMI can be broadly classified into three types

3.2.1 Hand-held portable transceivers (transmitter-receivers) with the antenna mounted directly on the transmitting unit, such as citizens band (CB) radios, walkie-talkies, security and fire or police transceivers, cellular telephones and other personal communication devices;

NOTE:

Some cellular telephones or similar devices transmit signals while they are ON, even though they are not in use.

3.2.2 Medium-range mobile transceivers used in police cars, fire trucks, ambulances and taxis usually having the antenna mounted on the outside of the vehicle; and

3.2.3 Long-range transmitters and transceivers, such as commercial broadcast transmitters (radio and TV broadcast antenna towers) and amateur (HAM) radios.

NOTE:

Other types of hand-held devices (cordless phones, laptop computers, AM/FM radios, TV sets, CD players, cassette players, and small appliances such as electric shavers and hair dryers, et cetera) so far as we know, are not likely to cause EMI problems.

3.3 Powered Vehicle Electromagnetic Interference (EMI)

Because EM energy rapidly becomes more intense as one moves closer to the transmitting antenna (source), the EM fields from hand-held radio wave sources (transceivers) are of special concern. It is possible to unintentionally bring high levels of EM energy too close to the powered vehicle's control system while using these devices. This can affect your powered vehicle's movement and braking. Therefore, the warnings listed below are recommended to prevent possible interference with the control system of the powered vehicle.

3.4 Warnings

Electromagnetic interference (EMI) from sources such as radio and TV stations, amateur radio (HAM) transmitters, two-way radios and cellular phones can affect powered vehicles and motorized wheelchairs.

Following the warnings listed below should reduce the chance of unintended brake release or powered vehicle movement which could result in serious injury.

3.4.1 Do not operate hand-held transceivers-receivers such as citizens band (CB) radios or turn ON personal communication devices such as cellular phones, while the powered vehicle is turned ON.

3.4.2 Be aware of nearby transmitters, such as radio or TV stations, and try to stay away from them.

3.4.3 If unintended movement or brake release occurs, turn the powered vehicle OFF as soon as it is safe.



CAUTION


- Some stores may have automatic (sensor) doors and alarm systems set at certain frequencies that might affect your powered wheelchair

3.4.4 Be aware that adding accessories/components or modifying the powered vehicle may make it more susceptible to EMI. There is no easy way to evaluate their effect on the overall immunity of the powered vehicle.


3.4.5 Report all incidents of unintended movement or brake release to your powered vehicle dealer or KARMA, and note whether there was a source of EMI nearby.

4. PARTS

KP-20.2 / 25.2



- | | |
|-----------------------------------|--------------------------------|
| 1. Push Hand | 9. Front Casters |
| 2. Backrest | 10. Pelvic Belt |
| 3. Armrest | 11. Controller |
| 4. Armrest Height Adjustment Knob | 12. Controller Connection Tube |
| 5. Side Panel | 13. Seat |
| 6. Brake | 14. Footrest |
| 7. Drive Wheels | 15. Calf Strap |
| 8. Motor | 16. Curb Climber (Optional) |




- 17. Back Joint (KP-25.2)
- 18. Armrest Release Base
- 19. Power Module
- 20. Freewheel Lever

- 21. Anti-tippers
- 22. Fuse
- 23. Battery

5. OPERATION

5.1 VRII Joystick Controller

- 5.1.1. Power ON/OFF
- 5.1.2. Joystick
- 5.1.3. Speed Adjustment Buttons
- 5.1.4. Battery & Diagnostics Indicator
- 5.1.5. Speed Indicator
- 5.1.6. Horn Button
- 5.1.7. Charging Socket



(Figure 5. 1)

5.2 How to Operate Your wheelchair

5.2.1 Power ON/OFF

Press the power button. Be sure not to touch the joystick while turning on the controller. Your battery & diagnostics indicator should light up. **(See Figure 5.1)**

5.2.2 Speed Adjustment Buttons

The speed adjustment buttons allows you to set various forward and reverse speeds. The settings are a percentage of maximum power.

(See Figure 5.1) Notice the orange dots on the speed indicator. The more orange dots, the faster your speed.



CAUTION

- Please slow down before you get used to control your powered wheelchair.
- Please slow down your speed when driving indoors.

5.2.3 Driving Speed

You can choose between 1 to 5 orange dots in the speed indicator, with 5 being the maximum speed. Please refer to "Max. Speed" in Section 10. SPECIFICATIONS.


Speed Scale	Speed Equivalencies and Recommended Circumstances
1-2	Speed equivalent to walking slowly on foot On an decline Driving indoors or in a narrow space
3-4	Equivalent to walking at a normal speed Driving outdoors on flat ground
5	Speed equivalent to a brisk walk Driving in an open space On an incline

5.2.4 Forward, Reverse and Braking

Push forward on the joystick to make the wheelchair move forward.

Pull back on the joystick for the wheelchair to move backward. Release the joystick and the wheelchair will quickly slow down, the electromagnetic brake will be activated to stop the wheelchair.

Come to a complete stop before turning the power OFF. **(Figure 5.2)**



(Figure 5. 2)



CAUTION

- The stopping distance will vary with your forward/reverse speed. Therefore, please gradually slow down and come to a complete stop well before any obstacles or danger.
- To park the wheelchair, be sure to park on flat ground and then turn the power to OFF.
- NEVER use the freewheel mode when going down an incline.
- The electromagnetic brake won't function in freewheel mode.

5.2.5 Horn Button

The horn will sound for the duration of time the button is pressed.


5.2.6 Battery Placement

Assure the lever of side mounting plate is released before you place the battery. Take the battery strap and put the battery into side mounting plate, then slide the battery to the end. Following the step as above to place the second battery, then connect the cable to controller module plug. Lock the lever of side mounting plate to finish the battery placement.

Note: The battery weight is 18 kg. Please carefully hold the battery during disassembly.

Controller position Adjustment

You can adjust your controller for convenience on control. Release the screw underneath the armrest with controller, adjust the proper position you demand and tighten the screw.



5.3 Shark Controller (Optional)

5.3.1. Power ON/OFF

5.3.2. Speed Adjustment Buttons


5.3.3. Horn Button

5.3.4. Joystick

5.3.5. Battery & Diagnostics Indicator

5.3.6. Speed Indicator

5.3.7. Charging Socket



((Figure 5. 3))

5.4. Power ON/OFF

Press the power button. Be sure not to touch the joystick while turning on the controller. Your battery & diagnostics indicator should light up. (See Figure 5.3)

5.5 Speed Adjustment Button

5.5.1 The speed adjustment button allows you to set various forward and reverse speeds.

5.5.2 The settings are a percentage of maximum power. (See Figure 5.3)

5.5.3 Notice the speed indicator. The more lights it shows, the faster you speed.



CAUTION


- Please slow down before you get used to control your powered wheelchair.
- Please slow down your speed when driving indoors.

5.6 Forward, Reverse and Braking

5.6.1 Push forward on the joystick to make the wheelchair move forward.

5.6.2 Pull back on the joystick for the wheelchair to move backward.

5.6.3 Release the joystick and the wheelchair will quickly slow down, the electromagnetic brake will be activated to stop the wheelchair.
(See Figure 5.4)



((Figure 5. 4))

5.7 Driving Speed

The speed scale is printed on the speed dial label, ranging from 1 to 5, with 5 being the maximum speed. Please refer to "Max. Speed" in Section 10.SPECIFICATIONS.

Speed Scale	Speed Equivalencies and Recommended Circumstances
1-2	⦿ Speed equivalent to walking slowly on foot ⦿ On an decline ⦿ Driving indoors or in a narrow space
3-4	⦿ Equivalent to walking at a normal speed ⦿ Driving outdoors on flat ground
5	⦿ Speed equivalent to a brisk walk ⦿ Driving in an open space ⦿ On an incline

5.8 To Stop

5.8.1 Release the joystick and the electromagnetic brake will stop the vehicle.

5.8.2 Come to a complete stop before turning the power OFF.



CAUTION

- The stopping distance will vary with your forward/reverse speed. Therefore, please gradually slow down and come to a complete stop well before any obstacles or danger.
- To park the wheelchair, be sure to park on flat ground and then turn the power to OFF.
- NEVER use the freewheel mode when going down an incline.
- The electromagnetic brake won't function in freewheel mode.

5.9 Horn Button

The horn will sound for the duration of time the button is pressed.

5.10 Battery Indicator

When the power is turned on, the battery indicator will light up with a possible total of three red, four orange and three green lines indicating the remaining power of the batteries. Both of the green light being lit indicates that the batteries are fully charged. If only the red lines are lit, this indicates that the batteries are (nearly out of or) out of power.



SUGGESTION

- It is recommended that you charge the batteries immediately when the battery indicator shows only three or less red lines.
- After charging or replacing new batteries, drive the wheelchair for 2-3 minutes to make sure the batteries are fully charged before going on a long journey.
- In wintertime, the batteries may respond slowly and the drive range may also be reduced. You must store your wheelchair indoors above freezing temperatures.
- When driving on an incline, the battery indicator light might move up and down. This is normal.
- Even if the batteries are used properly their capacity will decay over time, thereby reducing the drive range (maximum distance traveled per full charge). Thus, when the drive range becomes about 50% of what brand new batteries would offer, it's time to replace them with new batteries. Otherwise, the batteries may die unexpectedly leaving you in a possibly precarious situation.
- The drive range will be shortened when driving frequently on slopes or uneven ground, as this consumes more battery power.

5.11 Unfolding the Chair

- 5.11.1 Step 1. With the chair in front of you and facing away, take hold of both handles and pull up into position, making sure that the spring lever becomes locked.
- 5.11.2 Step 2. Take hold of both armrest and pull them apart until the backrest is almost completely open (about 30 cm).
- 5.11.3 Step 3. Place hands on the top of the seat support rails with fingers pointing inwards and push down until the seat is completely open and the seat rails are fully located in the retaining cups.

5.12 Folding the Wheelchair

- 5.12.1 Step 1. Put fingers under the middle of the seat at both front and rear, grasp and give a sharp pull upwards until the chair is nearly closed. Never grasp from the sides.
- 5.12.2 Step 2. Push the armrests together.
- 5.12.3 Step 3. Folding is completed by releasing the spring lever and folding back the backrest handles

5.13 Footrest Removal and Refitting

- 5.13.1 To remove footrest push / pull the release lever and rotate the footrest until it faces outwards.
- 5.13.2 To refit locate the footrest unit onto the frame lugs (at about 90 degrees to the side of the frame) and rotate it inward until the unit locks into place.

5.14 Armrest Flip-Back & Height Adjustment

5.14.1 You can flip your armrests back for convenience on movement.

Push the levers on the armrest release bases and flip the armrests back. When you put back the armrests, please make sure that the armrests are inserted into the armrest release bases correctly.

5.14.2 You also can adjust the height of the armrests for your comfort.

Release the armrest height adjustment knobs and adjust proper position you demand, and tighten the knob. Please align the knobs at the adjustment holes of the armrests. (See Figure 5.5)

5.15 Freewheel Levers

The KP-20.2/KP-25.2 is equipped with freewheel levers to allow for manual pushing of the powered wheelchair if required.

5.16 To Select Neutral from the Engaged Position

Rotate the two levers located on the top of both motors, 90 degrees outward. The wheelchair will now roll freely. (See Figure 5.6)

5.17 To Engage the Motors from the Neutral Position

Simply rotate the levers 90 degrees inward toward the attendant. (See Figure 5.7)



Figure 5. 5




Figure 5. 6




Figure 5. 7

5.18 Armrest Flip Up and Height adjustment

KP-20.2/KP-25.2 armrest is able to flip up for user to transfer. You may flip up the armrest by pressing the release lever located at the front base of the armrest. Assure the armrest is firmly locked into the position without movement after flip up. KP-20.2/KP-25.2 is also able to adjust the armrest height by following steps. (See Figure 5-8) Pull the knob into horizontal position to release the armrest. (See Figure 5-9) Adjust the armrest to your desired position and retighten the knob by pushing down the knob into vertical position to fix the armrest. (See Figure 5-10) °



Figure 5. 8



Figure 5. 9



Figure 5. 10



CAUTION

- Always ensure that the freewheel levers are engaged before turning on the controller.

5.19 Getting In And Out

Check control box is switched OFF when getting in and out.

5.20 Getting In

- 5.20.1 Step 1. Engage both parking brakes.
- 5.20.2 Step 2. Turn up the foot plates and swing footrest to the side.
- 5.20.3 Step 3. The user should now lower themselves into the seat using the armrests as supports.

5.20.4 Step 4. Turn the foot plates back down and re-locate the footrests so that the user can place their feet upon them.

5.20.5 Step 5. Release parking brake.

5.21 Getting Out

5.21.1 Step 1. Make sure that the chair is as close as possible to your transfer point.

5.21.2 Step 2. Back the chair sufficiently to ensure that the casters are pointing forward i.e. in line with the drive wheels and usually with the caster forks pointing forward. Ensure that where possible the rear of the chair is against a wall or, if not, it must be firmly held by an attendant.

5.21.3 Step 3. Engage both parking brakes.

5.21.4 Step 4. Turn up the foot plates and swing the footrests to the side.

5.21.5 Step 5. Slide forwards on the seat.

5.21.6 Step 6. Place one foot firmly on the ground with the knee at a right angle and the other foot slightly apart and further back.

5.21.7 Step 7. Place your hands on the front of the armrests and moving your weight forward, bring your head and shoulders up over your knees as you push yourself up.



CAUTION

- Never stand on the footplates when getting in to or out of the chair.
- Always ensure that the front casters are in line with the drive wheels and pointing forwards and that the back of the chair is secure either against a wall or firmly held by an attendant.

5.22 Motor Brakes

5.22.1 Electro-magnetic brakes operate automatically when the controller joystick is central with a delay of approximately 1/2 second. The controller automatically provides progressive deceleration before the electro-magnetic brakes come into operation.

5.22.2 To check the brakes are operating correctly: switch the wheelchair on and move the joystick in any direction. The brakes should disengage. Repeat this check in all four directions ensuring the brakes engage each time the joystick is released to its central position. An audible click from each motor indicates the brakes are working.

5.23 Parking Brakes - Hand Operated

5.23.1 The parking brakes operate directly on to the tyre providing positive locking of the drive wheels. They should only be used when the motors are disengaged. The brake levers are situated on the rear frame strut and can easily be operated by both user and attendant. Pull up for on & down for off. (ensure that the tyres are always inflated to the recommended pressure).

5.23.2 Care Points Avoid finger traps. Do not touch the drive wheels or front casters.

5.23.3 Prevent children and others from putting their hands near the drive wheels or front casters.

5.23.4 Never use the wheelchair with the drive wheels disengaged unless there is someone in attendance as the motor brake will not operate.

5.24 Install the Anti-Tippers

5.24.1 To Fit

Depress the rear button and insert it into the rear of frame.

The forward locating button must engage into the retaining hole in the frame.

5.24.2 To Detach

Depress the rear button and pull it out.



CAUTION

- KARMMA recommends that you should install anti-tippers in any occasion to prevent you tip over.

5.25 Transportation

5.25.1 KP-20.2 and KP-25.2 have been crash tested to ISO7176-19 and is suitable to use as a seat in vehicles that has been adapted for this purpose. The wheelchair must be securely restrained as described in this manual. KP-20.2 and KP-25.2 were tested with a four point webbing system. For further information on the use of wheelchair restraints, please refer to the manufacturer's user guide.

5.25.2 KARMMA always recommend that the user transfers to the vehicle seat, because using a wheelchair as a seat in a vehicle does not provide the equivalent level of safety of a vehicle seating system.



WARNING

- Please transfer to a vehicle seat in the vehicle in all permitted cases.

5.26 Before use



CAUTION

- If the back angle of the wheelchair is adjustable, do make sure that the wheelchair backrest should be positioned as close as possible to 90 degrees.

5.26.1 Please detach all easily removable accessories from the wheelchair and stow them in a safe place where they cannot move around and become a safety hazard.

5.26.2 A four point tie-down system should tightly fasten to the lower wheelchair frame marked with a yellow sticker as pictured.

5.26.3 An occupant restraint system using both a pelvis and harness must also be fastened to the transport vehicle to insure user safety. Please refer to the restraint system's instructions.



5.26.4 The attendant / transporter must double-check all fastening points for proper connection and tightness to assure the wheelchair users safety. The attendant / transporter must be fully cognisant of the fastening system's usage and its weight limitations.

5.26.5 See <http://www.unwin-safety.com/> for more information

5.27 Method of Restraint

5.27.1 KP-20.2 and KP-25.2 were tested with four point webbing systems, with the back in the upright position. Please refer to the restraint manufacturer' s user guide for further details of its operation.

5.27.2 Fitting the front restraint


5.27.2.1 Connect the straps around the wheelchair frame and caster bracket as pictured below. The anchor point is marked with a yellow sticker.

5.27.2.2 Take the brakes off and reverse the wheelchair to tension the front straps. Apply the wheelchair brakes (the brakes should be left on during transport)

5.27.3 Fitting the rear restraint

5.27.3.1 Attach rear straps into the rail approximately 300 mm behind the wheelchair backrest.

5.27.3.2 Attach the hooks to the wheelchair brackets at the rear of the wheelchair as pictured below. Tension the restraint until the wheelchair is secure.



Fitting the front restraint



Fitting the rear restraint


5.28 Fitting the Seat Belt

5.28.1 Fasten the three point safety belt around the user.

5.28.2 The occupied wheelchair should be located in a forward-facing configuration, with the parking brake applied.

5.28.3 Pelvic restraints should make full contact across the front of the body near the junction of the thigh and pelvis as pictured below (correct seat belt position).

5.28.4 Restraints should not be held away from body by wheelchair components such as armrest or wheels as pictured below (Incorrect seat belt position).



Correct seat belt position



Incorrect seat belt position



CAUTION

- When transporting or lifting, never hold on to attachments or accessories (front casters, drive wheels, handrims, footrests, armrests, anti-tippers etc.)
- All lifting should be done with a straight back. Straight legs to lift the chair. Carefully carry it up or down steps; be aware of how you place your feet, do not cross them and trip yourself up.

5.29 Adjusting the length of the pelvic belt

5.29.1 Identify the require length of the belt. (Figure 5-13)

5.29.2 Push the two part of belt through buckle to make the belt longer or shorter. (Figure 5-14)

5.29.3 Pull the two parts of belt tight. (Figure 5-15)

5.29.4 Push buckle up to the plastic clip so that the belt can not slip out of position. Please check that there is no gap between the clip and buckle. (Figure 5-16)



SUGGESTION

- We recommend that the belt is fitted quite tightly around the pelvis. Please tighten until you can just get your fingers between user and seat belt. Please check that the belt is fastened correctly as recommended. If this is not possible please consult your supplier.



Figure 5. 11



Figure 5. 12



Figure 5. 13



Figure 5. 14

5.30 How to Operate It

5.30.1 When you sit on the wheelchair, fasten the pelvic belt.

5.30.2 Adjust the belt and make it fit your body for comfort and security.

5.30.3 When you don' t use it, please fasten the belt to prevent the belt from dropping into drive wheels during motion.

5.31 Maintenance

5.31.1 The pelvic belt is made from 100% polyester fibre and the clasp is made from plastic. It does not require any maintenance but should be inspected every month to ensure that the edges are not frayed and weekly to make sure that it correctly fits the user.

5.31.2 If there are any signs of deterioration, please replace it immediately.

6. BATTERIES AND CHARGER

6.1 Charging the Batteries


Before you attempt charging, ensure you read and understand the instruction manual supplied with the battery charger, and you choose the correct input voltage for battery charger.

6.2 Charge Your Batteries If Any of the Following Conditions Occur

- 6.2.1 If you have used your wheelchair for more than an hour continuously.
- 6.2.2 You've used more than one third of the total charge.
- 6.2.3 When you haven't used the wheelchair for more than two weeks.
- 6.2.4 If the current charge is not enough to reach your final destination.

6.3 Be Sure To Follow the Procedures Listed Below Precisely

- 6.3.1 Turn the wheelchair OFF.
- 6.3.2 Connect the charger's cord to a power outlet.
- 6.3.3 Locate the charging socket at the front of the joystick housing.
Then connect the charger's round plug to the charging socket.
(See Figures 6.1 and 6.2)
- 6.3.4 The charger's orange LED will light up as soon as the charging starts. The charging time is at least 6 hours, but it could be as long as 10 hours, depending on the status of the batteries and temperature.
- 6.3.5 The orange LED will change to green when the charging is completed.
- 6.3.6 Disconnect the cord and pull out the round plug from the charging socket.



(Figure 6.1)



(Figure 6.2)



SUGGESTION

- Do not disconnect the charger cord if the charging is not completed. The battery life will be seriously shortened or decayed if the batteries are repeatedly used without being fully charged. Therefore, be sure to always charge the batteries fully.
- Always complete the charging by letting the charge LED light turn green. NEVER stop charging before it is complete.
- When finished charging, disconnect the charger plug from the charging socket as soon as possible. Even when the charger is turned OFF, the electric power will slowly discharge if the cord is not disconnected. DO NOT leave the charger ON connected to the wheelchair for more than 24 hours.
- Charging time depends on the surrounding temperature; it takes longer in the winter.



CAUTION

- Please follow the guidelines below to avoid accidents while charging.
- Be sure to only use a KARMA approved battery charger and charge the battery completely every time. The battery can be damaged if a non-genuine charger is used.
- Never disassemble or modify the charger. This will negate the warranty.
- Please charge in a well-ventilated area where the battery is not directly exposed to sunlight. Do not charge the battery under rainfall, in morning dew or in an area where the humidity is high.
- Do not cover the wheelchair with any waterproof cloth or other objects while charging. Don't charge in temperatures less than -10°C (14°F) or higher than 50°C(122°F) as the charger may not work well and the battery may become damaged.
- The charger is not guaranteed to be water resistant.
- Valve Regulated Lead-Acid batteries do not have a memory function like cell phone batteries have. Please get in the habit of recharging your VRLA batteries frequently.



WARNING

- Keep your wheelchair away from flammable objects while charging, as it may lead to fire or explosion of the battery.
- Since the battery may produce hydrogen, do not smoke while charging and charge in a well-ventilated area.
- Never connect or disconnect the plug (cord) with wet hands or if the plug (cord) is wet to avoid possible electric shock.
- Never attempt to recharge the batteries by attaching cables directly to the battery terminals or clamps. **DO NOT** jump start your wheelchair!
- **DO NOT** attempt to recharge the batteries while the powered wheelchair is in operation.



CAUTION

- New batteries **MUST** be fully charged prior to initial use of the powered wheelchair.
- Always charge new batteries before initial use or battery life will be reduced.
- As a general rule, recharge batteries as frequently as possible to assure the longest possible life and to minimize required charging time. Plan to recharge them when use of the powered wheelchair is not anticipated.

6.4 Charger

The battery charger supplied by KARMA is for indoor use only, it must be protected from moisture and external heat source.



CAUTION

- Before charging, make sure to set the 115V ↔ 230V switch to match the local AC voltage. If the switch is improperly set to a lower or higher voltage position, it will cause damage to the charger, and, Change the 115V ↔ 230V switch settings only when charger is disconnected from AC line voltage.



WARNING

- The fan inside the charger will start to work when you turn on the charger. (for 8A charger will start to work as temperature is over 50°C and stop running under 50°C) If you find that the fan does not function properly, **DO NOT** use the charger. Otherwise, the charger may overheat and start a fire.

6.5 Batteries

6.5.1 Do not expose the batteries to temperatures below -10°C(14°F) or above 50°C (122°F) when charging or storing the vehicle.

6.5.2 Exceeding the specified temperature range can lead to either freezing or overheating of the batteries. This will damage the batteries and shorten their lives

6.5.3 The batteries used on your wheelchair are maintenance free sealed lead acid batteries. Thus, it is not necessary to change or refill the battery liquid.



WARNING

- Do not open the battery seal cap at any time.
- After ANY adjustments, repair or service and BEFORE use, make sure that all hardware is tightened securely. Otherwise, injury or damage may occur.
- When removing/installing batteries, keep all foreign objects, especially metal, away from battery terminals.



CAUTION

- Place the powered wheelchair in a well-ventilated area where work can be performed without risking damage to carpeting or floor covering.
- Failure to use the correct battery size and/or voltage may cause damage to the powered wheelchair and give unsatisfactory performance. For replacement of batteries, please contact your local KARMA supplier/dealer in order to ensure that the new batteries are correctly be used.



WARNING

- There are many risks involved with improper disposal of batteries, whether damaged or not.
- Batteries may contain heavy metals and be considered hazardous waste.
- Many regions around the world now strictly enforce laws to help protect our environment. By not disposing of batteries properly, you may be liable for fines from your local through national government.
- By not recycling batteries, they end up in landfills that are already low on space. From these land fills, lead and other chemicals may leach into streams and rivers destroying ecosystems and poisoning local water

resources.

- Improper disposal of batteries may also cause injury to unsuspecting others that come in contact with them. If you come in contact with fluids or corroded powder from a battery, rinse with water and contact a physician if necessary. If contact with eyes occurs, rinse for 15 minutes and then visit a physician.
- Incinerating or burning a battery will cause it to explode! Please see the end of Section 11. Cleaning & Recycling for more information.
- Do not use an extension cord when charging your batteries. A risk of fire and /or electric shock could be encountered.

6.6 Cleaning the Batteries

If a battery is contaminated by dirty water, battery acid, dust or other soil, the battery will discharge quickly. Therefore, please follow the following steps to clean the battery:

6.6.1 Turn the power OFF ;

6.6.2 Use a clean cloth to wipe off the soiled part of the wheelchair around the battery.

6.6.3 Clean the terminals and inside of the battery clamps by using a battery cleaning tool, wire brush or medium grade sandpaper. Upon completion, these areas should be shiny, not dull.

6.6.4 Examine the battery clamps/terminals for corrosion, if the terminal is covered by white powder, please stop using the battery due to it indicates that there's possible leakage of the battery. Please replace the batteries. Carefully dust off all metal particles.

7. INSPECTION AND MAINTENANCE

7.1 Daily Check

Check the following items before driving. If you find anything abnormal, visit your KARMA wheelchair dealer for further inspection before using it.

Item	Inspection Content
Joystick Controller	Is the joystick loose and/or making noise? Can the joystick be turned left and right smoothly? Can the speed buttons be adjusted freely and do they function well? Does the wheelchair stop when the joystick is released completely? Does it light up when the power is ON? Is there enough power for your trip? Does the horn work well?
Motor	Is there any abnormal noise? Does the electromagnetic brake work properly?
Freewheel Mode	Do the freewheel levers work properly?
Driver's Seat	Does the driver's seat recline? Do the seat angle adjustment knobs(or buttons)work properly? Does the surface of the seat remain or break?
Tires	Are there any cracks or other damage to the tires? Check the tire tread depth and air pressure.
Other	Is there any abnormal motor noise?



CAUTION

- Visit a KARMA wheelchair dealer for inspection and maintenance if you find anything abnormal.

7.2 Regular Maintenance Record

In order to ensure that your wheelchair is in good condition, please contact KARMA authorized dealers regularly and make further wheelchair inspection and maintenance records. We recommend that you inspect and maintain your wheelchair **every six months**. A maintenance/service fee may apply.



SUGGESTION

- Even if you don't use the wheelchair for a long time, the wheelchair should still be maintained regularly.
- You should clean your wheelchair, inside and out, regularly to keep it beautiful and safe.

I - Inspection

A - Adjustment

△ - Repair

3 - Replacement

T - Tightening

L - Add Oil

Item/Period		1 mth	6 mth	12 mth	18 mth	24 mth	30 mth
Joystick Control Unit			I	I	I	I/3	I/3
Power Module			I	I	I	I/3	I/3
Controller Buttons			I	I	I	I/3	I/3
Connection to the Electrical System				I		I	I/A
Motor	Operation and Noise		I	I	I/A	I/A	I/3
	Electromagnetic Brakes		I	I	I	I/3	I/3
	Connection to the Electrical System			I		I	
Batteries	Loose or Corroded Terminals			I		I	
	Connection to the Electrical System			I		I	
	Damage to the Electrical Cord/Cable			I		I	
Charger	Charging Function and LED Light	I	I	I	I/△	I/△	I/△
	Socket, Plug and Cords			I		I	

Item/Period		1 mth	6 mth	12 mth	18 mth	24 mth	30 mth
	Fan Functions Properly (if applicable)			I		I/ ^Δ	
Driver's Seat	Recline Function			I/L		I/L	
Tires	Cracks/Surface Damage	I/3	I/3	I/3	I/3	I/3	I/3
	Tread Depth		I	I/3	I	I/3	I
	Abnormal Wear Pattern			I/3		I/3	
Wheel Rim	Tightness of Bolts and Nuts			T		T	
	Wear or Deformation Around Tires			I/3		I/3	

7.3 Tires

7.3.1 The condition of the tires is affected by the way you drive and use your wheelchair.

7.3.2 Inspecting Tire Tread: please check the tread depth regularly.
Replace the tires when the tread depth is less than 0.5 mm.



CAUTION

- Please ensure that the tire is flat before you disassemble the rim to replace the pneumatic tire.

7.4 General Maintenance

7.4.1 You must maintain the vehicle more frequently if you drive on grass, sand or gravel roads often.

7.4.2 Do not use water, oil or other chemical solutions to clean your wheelchair. Be sure NOT to spray the vehicle with water, as not to damage the electronic components and PCB circuit board.

7.4.3 Please clean the vehicle by wiping it with either a dry or moist cloth.

7.4.4 Please take the vehicle to an authorized KARMA wheelchair dealer for repairs and adjustments. Improper adjustments could lead to wheelchair malfunction or an accident.

7.4.5 Please use a soft and dry cloth to keep your wheelchair clean. Use moderate or mild detergent to clean the wheelchair, if necessary.



CAUTION

- Before performing any maintenance, be sure to turn the power OFF and unplug the charger if connected.
- DO NOT over-tighten hardware. This could cause damage to the frame and hardware.



SUGGESTION

- Do not directly spray water on your wheelchair to wash it, as this could lead to a malfunction in the electric system.
- Do not use gasoline, solvents or other abrasive solutions; the casing may become deformed or damaged.



WARNING

- After ANY adjustments, repair or service and BEFORE use, make sure that all hardware is tightened securely, otherwise injury or damage may occur.

7.5 Suggested Maintenance Procedures

- 7.5.1 Before using the powered wheelchair, make sure all nuts and bolts are tight.
- 7.5.2 Check all parts for damage or wear and have a qualified technician replace it if necessary.
- 7.5.3 Check all parts for proper adjustment.
- 7.5.4 Ensure that the batteries are properly maintained and charged as required.
- 7.5.5 Inspect brake for proper operation.
- 7.5.6 The wheels and tires should be checked periodically for cracks and wear, and should be replaced by a qualified technician.
- 7.5.7 Regularly check for loose nuts and/or bolts in the front casters and drive wheels. If loose, have them adjusted by a qualified technician.
- 7.5.8 Clean dirt, dust and grease from exposed components.
- 7.5.9 A qualified technician can be found with an authorized distributor/dealer.

7.6 Transporting and Storing

- 7.6.1 The wheelchair may be disassembled for ease of transport. Please refer to section 5 for more detailed information on the individual steps. Reverse the order for reassembly.
- 7.6.2 Please store the wheelchair in a location where it is out of direct sunlight, rain and dew.
- 7.6.3 When stored for a long time, please charge the batteries to full and then disconnect the battery terminals.
- 7.6.4 For more details, consult your KARMA wheelchair dealer.

8. OPTIONAL ACCESSORIES

8.1 Cushion

Reduces pressure, lowers temperature, fits any shape and prevents bedsores.

Suitable for: all wheelchair users, stroke patients, spinal cord injuries, office use and driving.



8.2 Curb Climber

Assist your chair to overcome 8 cm high obstacles

8.3 How to Install The Curb Climber

8.3.1 Insert the slot A to the screw. (See Figures 8.1, 8.2)


8.3.2 Press the lever B and put the slot C into the screw. (See Figures 8.3)

8.3.3 Finish installation. (See Figures 8.4)


8.4 How to Remove The Curb Climber

8.4.1 Press the lever B and remove the slot C from the screw.


8.4.2 Remove the slot A from the screw.



(Figure 8-1)



(Figure 8-2)



(Figure 8-3)



(Figure 8-4)

9. TROUBLESHOOTING

Trouble or Problems with Your Wheelchair

You can inspect the following before taking your wheelchair to a KARMA wheelchair dealer.


Problem	Checking Points and Corrective Action
My wheelchair does not start	Is the power turned OFF? Turn ON the power.
	Do the batteries still have power? Is the battery indicator light not on? Charge the batteries.
	Is the charger plugged into the wheelchair? Unplug the charger and put it away.

If you cannot solve the problem by yourself, contact your KARMA wheelchair dealer for help.

9.1 VRII Controller, Battery & Diagnostics Indicator

The following table indicates what the gauge will display for any given state.

Indication	Status	Meaning	What to do
	1 signal	The battery needs charging or there is a bad connection to the battery.	Check the connections to the battery. If the connections are good, try charging the battery.
	2 signals	The left hand motor has a bad connection.	Check the connections to the left hand motor.
	3 signals	The left hand motor has a short circuit to a battery connection.	Connect your service agent.
	4 signals	The right hand motor has a bad connection.	Check the connections to the left hand motor.

Indication	Status	Meaning	What to do
	5 signals	The left hand motor has a short circuit to a battery connection.	Connect your service agent.
	6 signals	The wheelchair is being prevented from driving by an external signal.	Please remove the charger.
	7 signals	A joystick fault is indicated.	Make sure that the joystick is in the center position.
	8 signals	A joystick fault is indicated.	Make sure that all connections are secure.
	9 signals	The parking brakes have a bad connection.	Check the parking brake and motor connections and make sure all connections are secure.
	10 signals	An excessive voltage has been applied to the control system.	Check the battery connections.
	10 signals + Speed indicator	A communication fault is indicated.	Make sure that the joystick cable is securely connected and not damaged.

Shark Controller (Optional), Battery & Diagnostics Indicator



The Battery Gauge is used to indicate power on and provides an estimate of the remaining battery capacity.

Every green LEDs lit indicate well charged batteries.

If only **amber and red** LEDs are lit, the batteries are moderately charged. Recharge before undertaking a long trip.

If **only red** LEDs are lit, the batteries are running out of charge.

Recharge as soon as possible.

The following table indicates what the gauge will display for any given state.

Display	Description	This means...	Notes
	ALL LEDs Off	Power is Off	
	ALL LEDs ON steady	Power is ON	Less LEDs imply a reduced battery charge.
	Left RED LED is flashing	Battery charge is low	The batteries should be charged as soon as possible.
	Right to left 'chase'	SHARK is being brought out of Lock mode	To unlock SHARK, Press the Horn button twice within 10 seconds.
	Left to right 'chase' alternating with steady display	SHARK is in programming, inhibit and/or charging mode	The steady LEDs indicate the current state of battery charge..
	ALL LEDs flashing slowly	SHARK has detected an Out Of Neutral At Power Up (OONAPU) condition	Release the joystick back to neutral.

Flash Codes



Flash codes indicate the nature of an abnormal condition directly from the SHARK Information Gauge. Without the use of any servicing tools, the condition can be simply diagnosed.

Number of Flashes	Description	
1	User Fault	Possible stall timeout or user error. Release the joystick to neutral and try again.
2	Battery Fault	Check the batteries and cabling. Try charging the batteries. Batteries may require replacing.
3	Left Motor Fault	Check the left motor, connections and cabling.
4	Right Motor Fault	Check the right motor, connections and cabling.
5	Left Park Brake Fault	Check the left park brake, connections and cabling.
6	Right Park Brake Fault	Check the right park brake, connections and cabling.
7	SHARK Control Unit Fault	Check the SHARK Communications Bus connections and wiring. Replace the Control Unit.
8	SHARK Power Module Fault	Check SHARK connections and wiring. Replace the Power Module.
9	SHARK Communications Fault	Check Battery voltage is greater than 17V. Check SHARK Bus Cable. Replace the SHARK Power Module. Replace the SHARK Control Unit.
10	Unknown Fault	Check all connections and wiring. Consult a service agent.
11	Incompatible Control Unit	The Control Unit is incompatible with the Power Module. Ensure the brand of the Power Module matches that of the Control Unit.

10. SPECIFICATIONS

Product Model	KP-20 Series		
Nominal seat width (mm/in)*	410 / 16	460 / 18	515 / 20
Nominal seat depth (mm/in)**	430 / 16	480 / 18	480 / 18
Front caster / Drive wheel size (in)	8 / 12	8 / 12	8 / 12
Backrest height (mm)	450 ~ 525	450 ~ 525	450 ~ 525
Headrest height above seat (mm)	N/A	N/A	N/A
Armrest to seat distance (mm)	225 ~ 285	225 ~ 285	225 ~ 285
Front location of armrest structure (mm)	245 ~ 500	245 ~ 500	245 ~ 500
Seat surface height at front edge (mm)	525	525	525
Footrest to seat distance (mm)	360 ~ 445	350 ~ 435	350 ~ 435
Effective seat width (mm)	430	480	530
Effective seat depth (mm)	470	515	515
Seat plane angle (X°)	6	6	6
Backrest angle (X°)	11	11	11
Seat to backrest angle (X°)	95	95	95
Leg to seat angle (X°)	110	110	110
Overall length with legrest (mm)	985 ~ 1115	985 ~ 1115	985 ~ 1115
Overall width (mm)	590	640	690
Overall height (mm)	960 ~ 1035	960 ~ 1035	960 ~ 1035
Folded length (mm)	755	755	755
Folded width (mm)	380	395	395
Folded height (mm)	910	910	910
Total mass (kg)	67.0	67.6	67.6
Total mass without batteries (kg)	35.7	36.3	36.3
Mass of the heaviest part (kg)	30.0	30.6	30.6
Maximum user weight (kg)	115 / 135 (pneumatic tires)		
Motor power (W)	200	200	200
Battery capacity (Ah)	50*2	50*2	50*2
Charger output current (A)	6	6	6
Drive range (km)***	32	32	32
Battery Compartment (L x W x H) (mm)	220 x 170 x 175 (single compartment)		
Maximum speed forward (km)	6	6	6

Minimum braking distance from max speed (mm)	1000	1000	1000
Obstacle climbing (mm)	50	50	50
Ground clearance (mm)	60	60	60
Minimum turning radius (mm)	780	955	1130
Minimum turning radius without user (mm)	650	825	1000
Reversing width (mm)	1180	1220	1260
Static stability downhill (X°)	16	16	16
Static stability uphill (X°)	14	14	14
Static stability sideways (X°)	16	16	16
Dynamic stability uphill (X°)	6	6	6
Ability to climb rated slope (X°)***	12	12	12

* **Nominal seat width** is measured the width of seat cushion at 120mm in front of the backrest. This value is mainly used in the product ordering process.

** **Nominal seat depth** is measured from the front end of the seat cushion to the back cushion. This value is mainly used in the product ordering process.

*** **Drive range/ Ability to climb rated slope** are estimated based on the following conditions: 20°C~35°C, 115 kg occupant, brand new fully charged batteries and driving on the flat road.

- The actual product specifications differ from the data in the table by ± 1 cm ± 0.5 kg.
- The actual product specifications may vary according to different configurations.
- KARMA reserves the right to modify information herein without further notice.
- If you want to know more about the measurement methods of wheelchair specifications, please scan the QR code to watch the instructions on the KARMA website.



11. CLEANING & RECYCLING

11.1 Cleaning

- 11.1.1 In order to lengthen the life of your wheelchair, it is suggested that you clean the wheelchair periodically (if not daily), especially after it is used in the rain or snow.
- 11.1.2 We suggest that you don't use your wheelchair in the rain or snow.
- 11.1.3 If you are traveling and get caught in the rain, we suggest you try to find shelter and turn OFF your wheelchair while waiting it out.
- 11.1.4 If you can't find shelter, proceed to your destination.
- 11.1.5 However we do not recommend you do this often as it may shorten the life of your wheelchair or cause damage.
- 11.1.6 Driver's seat upholstery: use soapy water. Wring out the sponge/cloth before scrubbing. Do not use the wheelchair until it is dry.
- 11.1.7 Frame: wipe with a damp cloth, but be sure not to splash/spray the wheelchair with water, as doing so may damage the motor, electrical wiring and battery.
- 11.1.8 If compressed air is available, you can use the compressed air to clean your wheelchair (do not exceed 7K and keep the nozzle at least 10 cm away from your wheelchair).
- 11.1.9 Casing: wipe with a damp cloth. After it is dry, coat the casing with a light layer of wax, if desired.

11.2 Post-Consumer Recycling

- 11.2.1 A post-consumer recycling plan should follow the Environmental Protection Administration's regulations.
- 11.2.2 The items listed below are for reference only as local law and regulation may vary.

Material	Part Name	Method
Aluminum-alloy	Frame	Recyclable
ABS	Various	Recyclable
Various	Battery	Recyclable
Various	Motor/Reduction Gear	Recyclable
Various	Electric Parts	Recyclable

11.2.3 There are many organizations in many countries you can find on the internet that will accept used goods for charity or disassemble electronic equipment and batteries to recycle them.

11.2.4 Please search the information highway to help less fortunate individuals and protect our planet for future generations.

11.2.5 If you do not have access to the world wide web please use your local phone directory, ask for operator assistance or go to your local government center for more information.

11.2.6 Various parts of this wheelchair should be recycled to the collection and recycling location nearest to you. Earth Day is every day.

Check with your local recycling center on where and how to recycle used batteries as this information varies from region to region.



12. WARRANTY

12.1 Serial Number

To ensure after-sale service and warranty, please record and fill in the serial number which can be found on the frame.

This warranty is VOID if the original chair serial number is removed or altered.

Model	KP-20.2/ KP-25.2
Serial Number	

12.2 Contents and Duration of Warranty

12.2.1 If the parts of the product need to be repaired or replaced due to manufacturing defects, they can be repaired or replaced free of charge within the warranty period (a repair of this kind will hereafter be called warranty maintenance).

12.2.2 Frame is warranted for a period of 5 years and electrical system (controller, motor, charger, actuator and PCB) for a period of 1 year after delivery to the customer, then the affected parts will be repaired or replaced free of charge.

12.2.3 Battery warranty is subject to the periods set by the manufacturers. Please contact your local dealer in detail.

12.2.4 Consumable parts such as bearings, cables, carbon brushes, circuit breakers, covers, footplates, light bulbs, pads, side guards, upholsteries, wheels and similar parts are excluded from the warranty.

12.2.5 Warranty maintenance can be performed by authorized KARMA wheelchair dealers only.

12.2.6 KARMA is not responsible for the transportation costs for repair and replacement.

12.2.7 Warranty maintenance is the repair or replacement according to the condition of the parts, and replaced parts will be the property of KARMA.

12.2.8 KARMA reserves the right to modify product specifications without further notice; KARMA is not responsible for any changes to product specifications.

12.3 The Claims from This Warranty Shall Not Arise for The Following Reasons

12.3.1 The product or part has been damaged by accident, improper use, incorrect assembly or poor maintenance. The evaluation of it is the privilege of KARMA.

12.3.2 Damage resulting from normal wear and tear.

12.3.3 Any overloading of the product, please check the label for maximum user weight.

12.3.4 The product or part has not been maintained or serviced in accordance with the manufacturer's recommendations as shown in the User Manuals and/or the Service Manuals.

12.3.5 Use or install parts or accessories which are not manufactured or not approved by KARMA.

12.3.6 Damage resulting from use in unsuitable environments.

12.3.7 Damage resulting from any addition, modification, or alteration of the product or parts without the approval by KARMA.

12.3.8 Damage resulting from external factors, such as chemicals, oil, debris, etc.

12.3.9 Damage caused by natural disasters.

12.4 The Responsibilities of the User

12.4.1 Fully adhere to operation guidelines set forth in this manual.

12.4.2 Perform self-inspection and basic maintenance.

12.4.3 Keep a record of self-inspections and basic maintenance.

12.4.4 Not permit any alteration, addition, or modification or use of parts not manufactured by KARMA.

12.4.5 Thoroughly read and understand this manual.

12.5 The Responsibilities of the Dealer

12.5.1 Follow this owner's manual and educate user about product operation and basic maintenance

12.5.2 Verify and check whether the warranty conditions are met.

12.5.3 Provide relevant warranty information and recommendations for improvements.

12.6 Transfer of Warranty Rights

If you purchase a second-hand KARMA wheelchair while it is still within the warranty period, bring this user manual and the wheelchair to the authorized dealer and perform necessary product inspection, maintenance and warranty inheritance procedure.

Full Name:	
Gender:	
Date of Birth:	
Address:	
Model:	
Serial Number:	
Date of Purchase:	
Purchaser Signature:	

Dealer's Data

Name of the Store:	
Phone No. and Address:	

If you have any suggestions on how to improve our products, please don't hesitate contacting your local dealer to let us know what you think of your wheelchair. Thank you and enjoy.

karma

A Better Wheelchair, A Better Fit



KARMA MOBILITY S.L European Support Center.
www.karmamobility.es
+34 984 390 907
karma@karmamobility.es

KARMA MOBILITY LTD. UK representative
www.karmamobility.co.uk
+44 845 630 3436
info@karmamobility.co.uk

KARMA MEDICAL PRODUCTS CO., LTD.
www.KarmaMedical.com
+886 5 206 6688
globalsales@karmam.com.tw

(110404050011) Release Date: Sep. 2022 V.20