


rehabhire&sales

YOUR
LEADING
HEALTHCARE
EQUIPMENT
PARTNER



Raz-AT600

User Manual



1300 000 030 | rehabhire.com.au

YOUR
LEADING
HEALTHCARE
EQUIPMENT
PARTNER

SAFETY PRECAUTIONS

Failure to comply may cause injury or damage!

It is important that the Raz Mobile Shower Commode Chair be properly assembled and that all adjustments be made carefully. Please ensure that the user(s), dealer(s), and attendant(s) have read this manual and thoroughly understood the safe operation of the Raz Shower Chair. Up-to-date manuals are available online at www.razdesigninc.com. For clarification of any section of this manual, contact your distributor.



CHAIR CLASSIFICATION BASED ON MAXIMUM USER WEIGHT	
AT600	275 kg (605 lbs)

SAFETY WARNINGS TO CONSIDER FOR USER PURPOSES

THIS DEVICE IS SUITABLE FOR TYPE A; HIGH USAGE (e.g. INSTITUTIONS) AND SHOULD NEVER BE USED FOR ANY PURPOSE OTHER THAN ORIGINALLY INTENDED

Stability

The casters on all models are equipped with dual locks. The step-on lever activates both the roll lock and the swivel lock. ***It is essential that all casters on the shower chair be in an outward position before engaging the locks, for optimal forward, rearward, and lateral stability.*** If so equipped, disengage the optional directional caster lock before rotating the rear caster to the rearward position and applying the caster's dual-locks.

USERS MUST NOT lean forward while sitting in the Raz Mobile Shower Commode Chair as this may cause the chair to tip forward.

USERS MUST NOT wheel down a ramp without an attendant who can hold onto the chair during the descent.



CASTER LOCKS ARE NOT BRAKES! Never use caster locks to slow down the shower chair when descending inclines as this could cause the chair to tip in the direction of travel.

Lateral Stability

USERS MUST NOT lean over the edge of a Raz Mobile Shower Commode Chair as this could cause the chair to tip.

USERS MUST NOT attempt to pick up objects from the floor by leaning over and reaching – use a reacher or ask for assistance from an attendant.

USERS MUST NOT position the Raz Mobile Shower Commode Chair on a side slope as this could cause it to tip.

Ramps and Other Inclines

Ramps and other surfaces that are not level (such as slopes into wheel-in showers) must be tested with an attendant to ensure they don't compromise the chair's stability. A qualified healthcare professional should be consulted to assist with this process.

DO NOT push or pull the Raz-AT600 up or down an incline or ramp that exceeds a 1:12 grade (5° slope).

DO NOT allow any user to be propelled (pushed or pulled) in a Raz-AT600 in any other manner than by an attendant.

ALWAYS CHECK THAT THE RAZ-AT600 IS STABLE BEFORE IT IS OCCUPIED BY ANY USER AT ANY TIME.

Immersion

The Raz Mobile Shower Commode Chair is not designed to be used as a pool chair. Do not immerse the shower chair in water or any other liquid, whole or in part. Immersion may result in damage to the shower chair.



Chlorine Exposure

Suitable for use in wet areas, but not pools and similarly corrosive environments.

DO NOT store the device near pools because chlorine in the air may cause rust on some of the chair's stainless steel components.

Transfers

The most dangerous part of using any shower chair is the transfer. Since every chair has different stability characteristics, it is essential to carefully review and test the transfer technique to ensure the procedure is safe. Test transfers should be performed with an attendant who can prevent the chair from tipping or moving during the transfer. If you feel that the transfers cannot be repeatedly performed safely, contact a healthcare professional who is familiar with transfer techniques and options.



Individuals who have not learned proper transfer skills must seek assistance during transfers and may require a mechanical lift.

Transferring onto a Raz-AT600


ALWAYS lock all four casters with front casters in the forward position for maximum chair stability.

Transfers should never be attempted with the shower chair positioned on slippery surfaces. It is recommended that transfers be performed on clean, dry, non-slippery surfaces. Hard surface flooring such as tile, wood, or concrete may not provide sufficient grip to keep locked casters and/or rear wheels from sliding. A fully carpeted surface is optimal. Be aware that area rugs and shower mats can slide on floors and thereby present a hazardous surface.



DO NOT stand on any part of the frame of the Raz-AT600.

NEVER use the footplates to assist in transfers. This could cause the chair to tip. The foot supports should be swung to the side or removed entirely during transfers to provide unobstructed access to the shower chair.



Tilt Function

ALWAYS check the tilt function and ensure both locking gas springs and release cable assemblies are fully operational **BEFORE** the chair is occupied by the user.

A pelvic strap can act as an additional safeguard for the user. A pelvic support strap may be ordered as an option with the Raz-AT600.

DO NOT attempt to tilt an occupant if the attendant is physically incapable of safely and smoothly operating the tilt function.

IMPORTANT !

USERS SHOULD NEVER BE LEFT UNATTENDED IN A RAZ MOBILE SHOWER COMMUNE CHAIR.


IMPORTANT !

DO NOT LIFT A RAZ MOBILE SHOWER COMMUNE CHAIR AND OCCUPANT BY THE ARM SUPPORTS OR FOOT SUPPORTS!

IMPORTANT !

ALWAYS BE SURE THAT THE SEAT IS PROPERLY INSTALLED AND LATCHED BEFORE A USER IS TRANSFERRED ONTO THE CHAIR. THE REAR SEAT ATTACHMENT BRACKETS MUST BE FULLY HOOKED UNDER THE REAR SEAT SUPPORT TUBES. THE FRONT SEAT ATTACHMENT BRACKETS MUST BE SECURELY LATCHED AROUND THE FRONT SEAT SUPPORT TUBES. BOTH FRONT AND REAR SEAT ATTACHMENT BRACKETS SHOULD BE SECURED IN PLACE BY TIGHTENING THE 8 SCREWS ON THE UNDERSIDE OF THE SEAT.

INSTRUCTIONS FOR THE PROPER REMOVAL AND INSTALLATION OF THE SEAT FOR CLEANING MUST BE PROVIDED TO THE USER AND / OR ATTENDANTS.



IMPORTANT !






THE RAZ-AT600 IS INTENDED FOR FOLLOWING USES:

- **AS AN AID FOR SHOWERING/BATHING, GOING TO THE TOILET, AND OTHER ACTIVITIES ASSOCIATED WITH PERSONAL CARE AND PERSONAL HYGIENE;**
- **AS A SUPPORT FOR THE BODY OF THE USER IN A SEATED OR RECLINED POSITION;**
- **WITH OR WITHOUT THE ASSISTANCE FROM A CAREGIVER;**
- **OVER A STANDARD TOILET, OR AS A COMMUNE WITH THE COMMUNE PAN;**
- **IN INDOOR ENVIRONMENTS ONLY;**
- **IN BOTH TYPE A, HIGH USAGE (e.g. INSTITUTIONAL) ENVIRONMENTS AND IN-HOME ENVIRONMENTS**

THE RAZ-AT600 DOES NOT PERFORM ANY THERAPEUTIC OR DIAGNOSTIC FUNCTIONS, AND IS NOT INTENDED FOR USE FOR ANY THERAPEUTIC OR DIAGNOSTIC PURPOSES. ALL PRODUCTS WITHIN THE RAZ FAMILY ARE NON-STERILE, REUSABLE, NON-INVASIVE, AND NON-IMPLANTABLE.

THE RAZ-AT600 IS INTENDED TO BE USED BY PEOPLE WITH MOBILITY IMPAIRMENTS, WHO MAY HAVE DIFFICULTY IN STANDING OR KEEPING THEIR BALANCE WHILE COMPLETING ACTIVITIES SUCH AS SHOWERING, TOILETING, OR PERSONAL HYGIENE. TYPICAL USERS HAVE LIMITED ENDURANCE AND POSTURAL CONTROL, AND MAY REQUIRE ASSISTANCE FROM A CAREGIVER FOR ACTIVITIES SUCH AS SHOWERING, TOILETING, OR PERSONAL HYGIENE. FOR PRODUCTS WITH RECLINE FUNCTION, USE OF THE RECLINE FUNCTION MAY BE CONTRA-INDICATED WHERE THE USER IS SENSITIVE TO SHEARING FORCES ON SEATED SURFACES WHEN OPENING THE SEAT-TO BACK-ANGLE.

LABELS OR SYMBOLS ON THE PRODUCT


PRODUCT DESIGNATION	RAZ-AT600
MANUFACTURER ADDRESS	 Raz Design Inc. 135 Railside Road Toronto, ON Canada M3A 1B7
AUTHORIZED REPRESENTATIVE DETAIL	<div style="border: 1px solid black; padding: 2px; display: inline-block; margin: 2px;">EC</div> <div style="border: 1px solid black; padding: 2px; display: inline-block; margin: 2px;">REP</div> UKRP <div style="border: 1px solid black; padding: 2px; display: inline-block; margin: 2px;">CH</div> <div style="border: 1px solid black; padding: 2px; display: inline-block; margin: 2px;">REP</div>
PRODUCT IDENTIFIER (PI)	<div style="border: 1px solid black; padding: 2px; display: inline-block; margin: 2px;">SN</div> BAT-XXXXXX
UDI-DI	<div style="border: 1px solid black; padding: 2px; display: inline-block; margin: 2px;">UDI</div> 628758003601
MEDICAL DEVICE	<div style="border: 1px solid black; padding: 2px; display: inline-block; margin: 2px;">MD</div>
EUROPEAN CONFORMITY	
DATE OF MANUFACTURE	
MAXIMUM USER WEIGHT	
READ USER MANUAL	

USER & DEALER MANUAL TABLE OF CONTENTS

1.0 GENERAL INFORMATION	
1.1 Raz-AT600 Mobile Shower Commode Chair Overview	8
1.2 Raz-AT600 Mobile Shower Commode Chair Specifications	8
2.0 SET-UP: INTENDED FOR DEALER USE ONLY	
2.1 Back Support.....	9
2.1.1 Back Support Installation	9
2.1.2 Back Support Depth Adjustment	9
2.1.3 Back Support Angle Adjustment	10
2.1.4 Adjustable-Tension Back Upholstery Adjustment.....	10
2.2 Caster Installation	10
2.3 Raz Seats	10
2.3.1 Raz Seat Installation.....	10
2.3.2 Ischial Pelvic Alignment System (IPAS) – Adjustment.....	11
2.4 Seat-To-Floor Height Adjustment	11
2.5 Weight Assisted Spring System (WASS) Adjustment	13
2.6 Adjustable Foot Supports	13
2.6.1 Foot Support Installation	13
2.6.2 Foot Support Adjustment	13
2.7 Molded Head Support	13
2.7.1 Head Support Installation.....	13
2.7.2 Head Support Adjustment.....	13
2.8 DEALER USE ONLY - Fastener Installation.....	14
3.0 USE AND OPERATION	
3.1 Tilt Function	15
3.2 Tilt-Assist Pedal	15
3.3 Flip-up Arm Supports	15
3.4 Adjustable-Tension Back Upholstery	16
3.5 Raz Seats	16
3.6 Adjustable Foot Supports (Footrests).....	17
3.7 Molded Head Support (Headrest).....	17
3.8 Commode Pan	17
3.9 Casters.....	18
4.0 CLEANING AND MAINTENANCE	
4.1 Cleaning After Each Use	19
4.1.1 Commode Pan.....	19
4.1.2 Seat.....	19
4.2 Cleaning As Needed	19
4.2.1 Adjustable-Tension Back Upholstery.....	19
4.2.2 Frame, Arm Support, Adjustable Foot Supports, and Other Components.....	19
4.3 Suggested Cleaning Products	19
5.0 LIMITED WARRANTY	
6.0 CONTACTS	

1.0 GENERAL INFORMATION

1.1 Raz-AT600 Mobile Shower Commode Chair Overview



1.2 Raz-AT600 Mobile Shower Commode Chair Specifications

SPECIFICATIONS	Raz-AT600
WEIGHT CAPACITY	275 kg (605 lbs)
FRAME WIDTH seat and back frame	56 cm (22")
OVERALL WIDTH	68 cm (26.75")
OVERALL LENGTH at 0° tilt, with foot supports	100 cm (39.5")
OVERALL LENGTH at 25° tilt, with foot supports	124 cm (48.75")
SEAT WIDTH commode seat	56 cm (22")
SEAT DEPTH front of the back post to the front of the seat	43-48 cm (17"-19")
CLEARANCE WIDTH	41 cm (16")
CLEARANCE HEIGHT 6" casters with pan brackets	46-48 cm (18"-19")
CLEARANCE HEIGHT 6" casters without pan brackets	48-50 cm (18.75"-19.75")
ARM SUPPORT (ARMREST) HEIGHT ABOVE SEAT	24 cm (9.5")
SEAT-TO-FLOOR HEIGHT with 6" casters	58-60 cm (22.75"-23.75")
PRODUCT WEIGHT with foot supports	38 kg (84 lbs)
FRAME AND HARDWARE MATERIAL	Stainless Steel (Lifetime warranty against corrosion)
MINIMUM TURNING CIRCLE	114 cm (45")

2.0 SET-UP: INTENDED FOR DEALER USE ONLY

IMPORTANT NOTE: If the threads are not lubricated, cold-welding of fasteners can occur.

- Apply a lubricant (liquid detergent, soap, lip balm, wax, etc.) to the threads before tightening the fasteners to reduce the possibility of cold welding.
- Do not overtighten fasteners.

IMPORTANT !

ADJUSTMENTS SHOULD NEVER BE ATTEMPTED WITH THE USER IN THE CHAIR.

2.1 Back Support

2.1.1 Back Support Installation

Tools required: 5mm hex key, 13mm wrench

1. Remove the Shower Chair from the shipping carton.
2. Lift the back support to an upright position.

TIP: Temporarily position the back support upright by inserting a Philips screwdriver through one plate and frame tube.

3. Insert an included button head cap screw through one of the two holes on the back support plate, then push the screw through the back support frame tube.


NOTE: Select the rear hole for a 90° back support angle or the front hole for a 97° back support angle.

4. Attach to the other end of the screw with the coved spacer, flat washer and locknut as shown below.


5. Tighten the assembly.

6. Repeat steps 3 to 5 for the other back support plate.

NOTE: Re-lubricate fasteners prior to tightening.




TOP VIEW OF THE BACK SUPPORT INSTALLATION



2.1.2 Back Support Depth Adjustment

Tools required: 5mm hex key, 13mm wrench

1. Loosen and remove the locknuts and flat washers on both inner sides of the seat frame.
2. Remove both screws from the back support mounting plates.
3. Reposition the holes of the back support mounting plates onto the holes of the seat frame for the desired seat depth.
4. With the back support in place, insert the screws into the holes of the back support mounting plates and push through.
5. Fasten the other end of the screws with washers and locknuts.
6. Tighten the assembly.




2.1.3 Back Support Angle Adjustment

Tools required: 5mm hex key, 13mm wrench

1. Loosen (do not remove) the locknuts, washers, and covered spacers on both vertical back support frame tubes.
2. Remove the lower screws and washers from the vertical back support frame tubes as shown in the image.
3. Rotate the back support to one of the two positions on the back support mounting plates.


NOTE: Select the front hole for a 97° back support angle or the rear hole for a 90° back support angle.

4. With the back support in place, insert the screws through the selected holes on the back support plates and push them through the frame tubes.
5. Fasten the other end of the screws with curved washers and locknuts.
6. Tighten the assembly with a hex key and wrench.



2.1.4 Adjustable-Tension Back Upholstery Adjustment

In order to adjust the back upholstery, each strap can be tightened or loosened in the buckles to achieve a wide range of positions. For example, the second and third straps from the bottom can be tightened to create more pronounced support in the lumbar area of the back.



2.2 Caster Installation


The Raz-AT600 chairs are shipped with casters not installed. Please follow the instructions below for caster installation.

Tool required: 5mm hex key

1. Insert caster stem into caster housing on the base frame.
2. Align the threaded hole in the caster stem with the hole in the caster housing.
3. Insert an M8 button-head screw into the hole on the caster housing.

NOTE: Ensure each M8 screw has some medium-strength thread locker solution on its threads before using.


4. Thread the M8 screw through the caster housing and into the caster stem.
5. Continue to thread the M8 screw until it is tight.
6. Repeat steps 1 to 5 until all four casters are secured in place.



2.3 Raz Seats

2.3.1 Raz Seat Installation

1. Ensure that the front spring brackets are in the open position. To open the front brackets, squeeze the tabs on the hinged portion of the bracket.
2. Slide the rear of the seat over the rear cross tubes as the front of the seat is lowered onto the frame. Ensure that the rear seat brackets hook below the cross tubes.
3. Continue lowering the front of the seat until the front seat brackets nestle onto the front seat support tubes.
4. Secure the seat in place by closing the front seat brackets. To close the seat bracket, squeeze the clip and close the bracket around the front support tubes.
5. Reverse steps 1 to 4 to remove the seat from the frame.



2.3.2 Ischial Pelvic Alignment System (IPAS) – Adjustment

The seat has approximately 5 cm (2") of fore/aft adjustment allowing optimal positioning of the seat aperture around the user's Ischial Tuberosities.


Tool Required: #2 Phillips screwdriver

1. Release the seat by opening the front seat brackets.
2. Lift the front of the seat and pull upward and then forward to remove the seat from the chair.
3. Loosen (but do not remove) the eight screws attaching the front seat brackets and rear seat brackets to the underside of the seat.

NOTE: For clarity, the rear seat brackets are shown below without commode pan brackets.

4. Place the seat back in position on the frame and close the front seat brackets around the front seat support tubes.
5. Slide/wiggle the seat to the desired position.
6. Ensure rear seat brackets are firmly positioned against the rear seat support tube.
7. Tighten all eight screws attaching the front and rear seat brackets to the seat to prevent further movement of the seat.

NOTE: Ensure the front and rear seat brackets are securely latched around the front and rear seat support tubes before allowing a user to transfer into the chair.






2.4 Seat-To-Floor Height Adjustment

To change the seat-to-floor height on a Raz-AT600 model, adjust the pivot post heights on both sides of the chair.

Tools required: 5mm hex key, 13mm wrench

1. Ensure the chair is out of tilt by squeezing both tilt triggers.
2. Remove the seat from the chair; squeeze the front seat brackets to open them, lift the front of the seat, and pull the seat out of the chair (refer to 2.3.1).
3. Remove the foot supports if they are installed.
4. On a soft and clean surface (e.g. carpeted surface, flattened packaging box), carefully tip the chair forward as in the drawing below.







5. Remove the screws and locknuts ① from both gas strut extension tubes.
6. Remove the locknuts and washers ② from both pivot posts.
7. Remove the screws ③ from both sides.

NOTE: Use a hex key to back out the screws, the screws are threaded into the pivot posts.

8. Move the base frame along the pivot posts to the desired seat-to-floor height.
9. Align the pivot posts' holes to the base frame's holes.
10. Insert the screws ③ into the aligned holes on both sides.

Tip: Insert one screw on each side first rather than inserting both screws on one side before the other side.

11. Attach the locknuts ②.



12. Loosen (do not remove) the screws and locknuts ④ on both spring support blocks.
13. Level the seat frame to be parallel with the base frame, and align the eyelet of each gas strut to the nearest hole of the gas strut extension tube.
14. Install the screws through the aligned holes and attach the locknuts ①. The screws must be inserted through the gas strut extension tube and the eyelet of the gas strut as shown.
15. Adjust both spring support blocks so that the spring bushing is in contact with the spring support tube, and tighten the screws and locknuts ④.
16. Ensure all hardware is installed securely before raising the chair to an upright position.

IMPORTANT !

INCREASING THE SEAT HEIGHT AND THE BACK SUPPORT ANGLE WILL DECREASE REARWARD STABILITY. IT IS ESSENTIAL THAT A STABILITY CHECK BE PERFORMED AFTER ANY OF THESE ADJUSTMENTS ARE MADE.


2.5 Weight Assisted Spring System (WASS) Adjustment

Tool required: 30mm wrench

The engagement of the WASS can be altered by adjusting the hex nuts along the length of the rod to help facilitate the tilting action for the attendant depending on user weight.

Move the hex nuts up the rod to engage the springs further. This is recommended for heavier users as the springs will help support the user's weight when tilting. Move the hex nuts down the rod to partially disengage the springs if the tilting action poses difficulty for the attendant.


When more adjustment is needed, adjust the spring support block along the base frame using a 4mm hex key and a 10mm wrench.



2.6 Adjustable Foot Supports


2.6.1 Foot Support Installation

1. Remove the foot supports from the inner accessory box.
2. Place the foot support hanger tubes into the receivers, welded at the front of the chair frame. *If the support tubes are directed forward, the pins on the tubes will engage with grooves in the receivers and lock in place.*



2.6.2 Foot Support Adjustment

1. Remove a plastic snap-pin from one foot support hanger tube.
2. Adjust the footplate extension tube vertically to achieve the desired height in the foot support hanger tube.
3. Insert the snap-pin through the foot support hanger tube and through the extension tube. Ensure that the snap-pin is completely through the tube and that the green plastic clip is snapped securely around the outside of the hanger tube.
4. Repeat steps 1 to 3 for the other footplate.



2.7 Molded Head Support

A molded head support is included with all Raz-AT Shower Chairs. It is shaped to cradle the head comfortably and is easy to clean. The head support can be adjusted indefinitely forward and backward within a 5" range and vertically within an 11½" range.


2.7.1 Head Support Installation

1. Remove the head support assembly from the inner accessory box.
2. Insert the square tube of the head support assembly into the receiver on the back support while ensuring the molded pad is facing forward.
3. Unfold the head support by loosening the Tri-lobes.
4. Adjust the angle and height to the user's preference.
5. Tighten the Tri-lobes fully to secure the angle and height of the molded head support.

2.7.2 Head Support Adjustment

To adjust the position of the head support:


1. Loosen the Tri-lobes on the head support assembly.
2. Adjust the angle and height to the user's preference.
3. Tighten the Tri-lobes to fully secure the head support position.



To adjust the angle of the head support pad:


Tool required: 4mm hex key

1. Loosen (do not remove) the three screws on the back of the molded head support.
2. Rotate the head support pad around the end of the head support mounting bar to suit the user's needs.
3. Tighten the screws to secure the position of the pad.
4. Ensure that all knobs and screws are tightened securely for the user's safety.



2.8 DEALER USE ONLY - Fastener Installation

All fasteners should be installed securely. Locknuts removed for any reason may need to be replaced rather than re-used. Check the performance of nyloc nuts when re-attaching to the screw. There should be significant detectable resistance when tightening. If not, replace the nyloc nut with a new one. It is advisable to check annually that all fasteners are secure.




3.0 USE AND OPERATION

3.1 Tilt Function

The Raz-AT600 tilts up to 25° to provide comfort and relieve pressure for the user. Two gas spring struts facilitate this tilting action, making the tilt smooth for the user and the attendant. The attendant must squeeze both tilt triggers to tilt the chair to the desired angle. To lock the chair in place, the attendant can release the triggers. Directly after purchase, the triggers and springs may seem stiff. This stiffness is normal and diminishes with use. Using the tilt-assist pedal will help to tilt the seat frame back. The tilt function must only take place when the casters are locked.

NOTE: When the chair is mounted at the lowest height, the maximum potential tilt angle is 25°. The higher the chair is set, the fewer degrees the chair can be tilted back.




IMPORTANT !

DO NOT ATTEMPT TO TILT A RAZ SHOWER CHAIR IF THE ATTENDANT IS PHYSICALLY INCAPABLE OF SAFELY AND SMOOTHLY OPERATING THE TILT MECHANISM WITH A USER OCCUPYING THE CHAIR. ANY ATTENDANT MUST BE FAMILIAR WITH PROPER TILTING TECHNIQUES PRIOR TO ASSISTING A USER IN THE SHOWER COMMODE. A PELVIC STRAP IS RECOMMENDED FOR ANY USER AS AN EXTRA SAFETY PRECAUTION.


3.2 Tilt-Assist Pedal

A Tilt-Assist Pedal is standard on the Raz-AT600 and is used to assist in tilting the chair for attendants.



3.3 Flip-up Arm Supports

The arm supports can flip up to allow for ease of transfer, and lock automatically (if the optional lock is included) when flipped down.




3.4 Adjustable-Tension Back Upholstery


The Multi-Position Back Support comes with an Adjustable-Tension Back Upholstery that provides trunk support for optimal positioning and comfort. The upholstery is removable for machine washing.

To install the Adjustable Tension Back Upholstery:

1. Place the fabric against the front of the Back Support canes making sure that when wrapped around, the clips on the straps face away from the fabric and the loops point downwards.
2. The first (1) and second (2) straps are positioned above the crossbar.
3. The third (3), fourth (4), and fifth (5) straps are positioned between the crossbar and armrest mounting plates.
4. The sixth (6) strap is positioned below the armrest mounting plates.
5. Adjust all of the straps to meet the user's needs.



In order to adjust the back upholstery, each strap can be tightened or loosened in the buckles to achieve a wide range of positions. For example, the second and third straps from the bottom can be tightened to create more pronounced support in the lumbar area of the back.




To remove the Adjustable-Tension Back Upholstery, release the buckles on the straps.

3.5 Raz Seats

To install a Raz Seat:

1. Ensure that the front spring brackets are in the open position. To open the front brackets, squeeze to unlatch.
2. Slide the rear of the seat over the rear cross tubes as the front of the seat is lowered onto the frame. Ensure that the rear seat brackets hook below the cross tubes.
3. Continue lowering the front of the seat until the front seat brackets nestle onto the front seat support tubes.
4. Secure the seat in place by closing the front seat brackets. To close the seat bracket, squeeze the clip and close the bracket around the front support tubes.
5. Reverse steps 1 to 4 to remove the seat from the frame.



IMPORTANT !

NEVER ATTEMPT TO CHANGE THE SEAT HEIGHT WITHOUT THE HELP OF A QUALIFIED DEALER. HIGHER SEAT POSITIONS WILL REDUCE THE CHAIR'S STABILITY AND MAY RESULT IN TIPPING AND INJURY.

3.6 Adjustable Foot Supports (Footrests)


Adjustable Footrests are standard on all Raz Mobile Shower Commode Chairs. The footrests can be flipped up, rotated away, or removed entirely for transfers. Adjustable Footrests can be extended for users with longer legs with the optional footrest extension tube, a 2" longer tube that can be ordered for any Raz Mobile Shower Commode Chairs.

To insert the Adjustable Footrests:

1. Place the footrest hanger tubes into the footrest receivers.
2. If the hanger tubes are aligned forward, the pins on the tubes will engage with grooves in the receivers and lock in place.

To adjust the length of the Adjustable Footrests:

1. Remove the snap-pin from one footrest hanger tube.
2. Reposition the footplate to a length suitable for the user.
3. Insert the snap-pin through the footrest hanger tube and through the inner tube.



To rotate the Adjustable Footrests:

1. Lift the footrest hanger tubes to disengage the pins.
2. Rotate the footrests outward or inward.


To remove the Adjustable Footrests, lift the footrests out of the receiving sleeves on the frame.

3.7 Molded Head Support (Headrest)

All Raz-AT600 include a molded headrest. It is shaped to cradle the head comfortably and is easy to clean. It can be adjusted indefinitely forward and backward within a 5" range and vertically within an 11½" range.

To adjust the molded headrest:


1. Loosen the Tri-lobes on the molded headrest assembly.
2. Adjust the angle and height to the user's preference.
3. Tighten the Tri-lobes fully to secure the angle and height of the molded headrest.
4. Ensure that all Tri-lobes are tightened securely for the user's safety.



3.8 Commode Pan

The Raz Commode Pan is designed with an elongated shape that fits directly under the seat to minimize backslash. A unique splash-reduction ridge is located at the center of the pan. The high gloss surface facilitates cleaning and rinsing, while the pouring spout and handle facilitate emptying. Removal of the Commode Pan is achieved by sliding the pan rearward.

NOTE: When re-inserting the pan, ensure that the pan is guided into the green plastic brackets and pushed forward until it hits the stops.



3.9 Casters

Each Raz Mobile Shower Commode Chair caster is equipped with a dual-lock feature, which stops the caster from swiveling and rolling simultaneously. To engage the dual-lock, step on the end of the caster brake lever until it latches in the “on” position. To deactivate the dual-lock, step on the raised end of the lever until it releases.



ALWAYS release the directional lock (if applicable) and swivel all four casters to their outward position for maximum stability before transferring or tilting a user.

IMPORTANT !

A QUALIFIED RAZ DEALER SHOULD BE CONTACTED TO CHECK ALL FASTENERS, CASTERS, HARDWARE, AND ALL OTHER MOVING COMPONENTS AT LEAST ONCE PER YEAR.

4.0 CLEANING AND MAINTENANCE

4.1 Cleaning After Each Use

4.1.1 Commode Pan

1. Remove the commode pan and empty the contents.
2. Wash the pan with a multi-purpose disinfectant cleaning solution. Use a soft brush if needed.
3. Rinse well before storage or re-use.

NOTE: A soapy hot water wash followed by a disinfectant spray can be substituted for a multi-purpose disinfecting cleaning solution.

4.1.2 Seat

1. Spray with a multi-purpose disinfectant cleaner and scrub with a soft brush as required.
2. Rinse with a water spray before storage or re-use.
DO NOT IMMERSE the seat in water or cleaning solutions.
3. For a more thorough cleaning, remove the seat from the shower chair to allow better access to the seat's underside and around the seat's opening.

4.2 Cleaning As Needed

4.2.1 Adjustable-Tension Back Upholstery

1. Remove the plastic buckles and tri-glides from the back upholstery straps, paying attention to how the straps are wound through each component as they are removed.
2. Remove the back upholstery from the back frame and machine wash with a mild detergent in warm water on a normal or gentle wash cycle. Include other strap accessories such as the H-strap, and pelvic and chest positioning straps.
3. Install the back upholstery and other positioning accessories on the chair and allow it to dry.
DO NOT dry the back upholstery in a clothes dryer or damage to the upholstery may occur.
4. Assemble all plastic buckles and tri-glides onto straps.
5. Adjust the straps to the desired tension.

4.2.2 Frame, Arm Support, Adjustable Foot Supports, and Other Components

1. Spray the frame and components with a multi-purpose disinfectant cleaner, scrub with a soft brush, and rinse with water.
2. If necessary, spray the frame and components with a diluted (10 parts water to 1 part solution) calcium/lime remover (such as CLR), scrub with a soft brush, and rinse thoroughly with water.
DO NOT mix a calcium/lime removal solution with any other liquid or toxic fumes may occur.
DO NOT use chlorine bleach on any Raz components.

It is recommended to the user that a qualified Raz dealer check over all fasteners and subassemblies if there is cause for concern regarding the shower chair's performance.

Several times each year the shower chair and its components should be wiped down with a diluted calcium/lime removal solution to reduce deposits, followed by a hot water rinse.

To reduce the risk of mildew buildup, the entire chair should be wiped down with a soft dry cloth after the final rinse and before storage or use.

4.3 Suggested Cleaning Products

Task	Recommended Products
General Cleaning	Antibacterial Detergent
Tuberculocidal Surface Cleaner Intermediate Level Disinfectant	Oxiver TB (Virox Technologies Inc.) SepteFX 7D-11 (Titan Worldwide Inc.)
Calcium and Lime Deterrent	CLR

IMPORTANT !

DO NOT USE ABRASIVE SCRUBBING POWDERS OR SCOURING PADS ON THE RAZ SHOWER COMMODE CHAIR.

DO NOT SUBMERGE THE RAZ SHOWER CHAIR IN WATER.

RINSE THOROUGHLY AFTER CLEANING TO ENSURE ANY SOAP OR DETERGENT RESIDUE IS REMOVED.

DO NOT USE CLEANING PRODUCTS WITHOUT CONSULTING THE PRODUCTS' INSTRUCTIONS

5.0 LIMITED WARRANTY

This warranty is extended only to the original purchaser/user for Raz Mobile Shower Commode Chairs that have been purchased from a qualified Raz dealer.

This warranty gives you specific legal rights, but you may have additional legal rights depending on your region and country.

This warranty shall be extended to comply with state/provincial laws and requirements.

With regards to the original purchaser/user only, Raz Design warrants its base and back frame components to be free of defects in materials and workmanship for a period of two (2) years from the date of purchase. All other components on the Raz Mobile Shower Commode Chairs are covered for a period of one (1) year and are subject to normal wear and tear exclusions. If within such warranty period any such product shall be proven to be defective, the part will be repaired or replaced, at Raz Design's discretion. This warranty does not include any labor or shipping charges incurred in replacement, part installation, or repair of any such product.

For warranty service, please contact the dealer from whom you purchased your Raz Mobile Shower Commode Chair. In the event that you do not receive satisfactory warranty service, please contact Raz Design directly using the contact information located at either the beginning or the end of this manual. In your correspondence, please provide your dealer's name and address, the product type, the date of purchase, the chair's serial number (on a grey label located on the chair frame), as well as the nature of the product issue. Do not return products to the manufacturer without prior authorization.

Limitations and Exclusions:

The foregoing warranty shall not apply to serial numbered products if the serial number has been removed or defaced, and to products subjected to negligence or an accident. Improper operation, maintenance, or storage may void the warranty coverage on the Raz Mobile Shower Commode Chair.

The warranty does not apply to products modified without Raz Design's express written consent. This includes but is not limited to, modification through the use of unauthorized parts or attachments, products damaged by reason of repairs made to any component by an individual or individuals who are not employed by a qualified Raz dealer, or to a product damaged by circumstances beyond Raz Design's control. Only Raz Design will have the authority to determine whether or not the circumstances are eligible for warranty coverage. The warranty will not apply to problems arising from normal wear or failure to adhere to the instructions contained in this manual.

The foregoing warranty is exclusive and in lieu of any other warranties, whether express or implied, including the implied warranties of merchantability and fitness for a particular purpose. The sole remedy for violations of any warranty whatsoever will be limited to the repair or replacement of the defective product in compliance with the terms contained in this document.

The application of any implied warranty shall not extend beyond the duration of the express warranty provided herein. Raz Design will not be liable for any consequential or incidental damages to part or whole of a Raz Mobile Shower Commode Chair.

Raz Design Inc. frame and hardware material are made of stainless steel for which we offer a lifetime warranty against corrosion. Please be advised this lifetime warranty is only extended to the original purchaser/user of Raz Shower Chairs that have been purchased from a qualified Raz dealer. Failure to comply with proper cleaning instructions may cause corrosion to the frame thereby voiding this warranty. Additionally, all claims of corrosion are subject to review and a request may be made for the return of the device in question, for a thorough examination, before a replacement order can be processed.

6.0 CONTACTS

It is important that the Raz Mobile Shower Commode Chair be properly assembled and all adjustments be made carefully. Up-to-date manuals can be downloaded from www.razdesigninc.com. For clarification of any section of this manual, contact your distributor or the manufacturer.



IMPORTANT !

PRODUCTS THAT CAN NO LONGER BE USED ARE TO BE DISPOSED OF SEPARATELY FROM HOUSEHOLD WASTE. THIS MUST BE DONE ACCORDING TO THE LOCAL AND NATIONAL REGULATIONS FOR ENVIRONMENTAL PROTECTION AND RAW MATERIAL RECYCLING. PLEASE LOOK AT YOUR LOCAL GOVERNMENT WEBSITE FOR FURTHER INFORMATION ON THE CORRECT DISPOSAL.

IF IT IS EVER FELT THAT ANY ASPECT OF THE RAZ SHOWER CHAIR IS IN NEED OF MAINTENANCE, CONTACT YOUR DEALER. ANY PROBLEMS THAT COULD AFFECT THE SAFE USAGE OF THE CHAIR MUST BE ATTENDED TO IMMEDIATELY. THE DEALER THAT PROVIDED THE SHOWER CHAIR MUST BE CONTACTED SO THAT REPAIRS CAN BE MADE PRIOR TO RESUMING THE USE OF THE CHAIR. IT MUST BE MADE CLEAR TO THE DEALER THAT THE SHOWER CHAIR IS IN NEED OF MAINTENANCE AND IS UNAVAILABLE FOR USE UNTIL REPAIRS HAVE BEEN UNDERTAKEN AND ALL PROBLEMS RESOLVED.



Notes

Lined area for notes, consisting of 25 horizontal lines.

rehabhire & sales

1 300 000 030

320 Lorimer Street Port Melbourne
contact@rehabhire.com.au
rehabhire.com.au

rehabhire & sales **REPAIRS AND SERVICE**

REGULAR MAINTENANCE OR REPAIRS

1 300 000 030

Unit 6 | 52 Wirraway Drive Port Melbourne
service@rehabhire.com.au
rehabhire.com.au/repairs-and-service

rehabinstallation

CAN MODIFY EVERY ROOM IN YOUR HOME TO
ENSURE IT IS ACCESSIBLE, SAFE, AND COMFORTABLE

1 300 000 030

320 Lorimer Street Port Melbourne
contact@rehabinstallation.com.au
rehabinstallation.com.au



Registered NDIS Provider